

Deep learning in photoacoustic imaging: a review

Handi Deng,^a Hui Qiao^{b,c,d,e}, Qionghai Dai,^{b,c,d,e} and Cheng Ma^{a,f,*}

^aTsinghua University, Department of Electronic Engineering, Haidian, Beijing, China

^bTsinghua University, Department of Automation, Haidian, Beijing, China

^cTsinghua University, Institute for Brain and Cognitive Science, Beijing, China

^dTsinghua University, Beijing Laboratory of Brain and Cognitive Intelligence, Beijing, China

^eTsinghua University, Beijing Key Laboratory of Multi-Dimension and Multi-Scale Computational Photography, Beijing, China

^fBeijing Innovation Center for Future Chip, Beijing, China

Abstract

Significance: Photoacoustic (PA) imaging can provide structural, functional, and molecular information for preclinical and clinical studies. For PA imaging (PAI), non-ideal signal detection deteriorates image quality, and quantitative PAI (QPAI) remains challenging due to the unknown light fluence spectra in deep tissue. In recent years, deep learning (DL) has shown outstanding performance when implemented in PAI, with applications in image reconstruction, quantification, and understanding.

Aim: We provide (i) a comprehensive overview of the DL techniques that have been applied in PAI, (ii) references for designing DL models for various PAI tasks, and (iii) a summary of the future challenges and opportunities.

Approach: Papers published before November 2020 in the area of applying DL in PAI were reviewed. We categorized them into three types: image understanding, reconstruction of the initial pressure distribution, and QPAI.

Results: When applied in PAI, DL can effectively process images, improve reconstruction quality, fuse information, and assist quantitative analysis.

Conclusion: DL has become a powerful tool in PAI. With the development of DL theory and technology, it will continue to boost the performance and facilitate the clinical translation of PAI.

© The Authors. Published by SPIE under a Creative Commons Attribution 4.0 Unported License. Distribution or reproduction of this work in whole or in part requires full attribution of the original publication, including its DOI. [DOI: [10.1117/1.JBO.26.4.040901](https://doi.org/10.1117/1.JBO.26.4.040901)]

Keywords: photoacoustic imaging; deep learning; convolutional neural network.

Paper 200374VRR received Nov. 18, 2020; accepted for publication Mar. 18, 2021; published online Apr. 9, 2021.

1 Photoacoustic Imaging

1.1 Brief Introduction to Photoacoustic Imaging

Photoacoustic imaging (PAI), also referred to as optoacoustic imaging, is an emerging imaging technique that works in both the optically ballistic and diffusive regimes. In addition to providing good contrast of blood vessels,¹ PAI is also capable of functional (such as blood oxygen saturation, sO_2) and molecular imaging.^{2,3}

PAI relies on the photoacoustic (PA) effect, which was first discovered by Alexander Graham Bell in 1880 as a conversion of light intensity modulation into sound emission.⁴ The rapid development of PAI in the past 30 years was incentivized by the maturation of laser and ultrasound (US) technologies. In a common PAI setting, when biological tissue is illuminated by a short optical pulse, a local temperature rise is induced, which in turn produces ultrasonic waves. Acoustic detectors outside the tissue receive the PA signal and an image is subsequently

*Address all correspondence to Cheng Ma, cheng_ma@tsinghua.edu.cn

reconstructed digitally to visualize the initial pressure rise distribution (which is closely related to the optical absorption distribution).

1.2 Image Formation

PAI can be categorized into two types according to the image formation principle. The first is photoacoustic microscopy (PAM),⁵ which is based on raster-scanning a focused ultrasonic transducer or a beam of focused light, to acquire an image pixel by pixel (each pixel is an A-line in the longitudinal direction; resolution along the A-line is generated by acoustic delay). In this implementation, the received PA signal mainly comes from the focal line, thus all spatial information is unmixed and no reconstruction is needed. The second type is photoacoustic computed tomography (PACT),⁶ in which signals are detected by a multi-element transducer array (or by scanning a single unfocused transducer). Each transducer element has a large acceptance angle within the field of view, and a PA image can be reconstructed by merging the data from all transducer elements. Popular arrays used in PACT include linear,^{7,8} ring (and partial ring),^{9,10} spherical (and partial spherical), and planar arrays.^{11,12} Linear arrays are easy to operate and relatively cheap, thus they were widely used for clinical applications. However, their angular acceptance is limited, resulting in poor image quality. In contrast, ring and spherical arrays have wider signal acceptance, and typically they generate better images. Yet, due to geometrical confinement, these arrays are often used for breast imaging and animal imaging and are relatively expensive. In all cases, the US transducers are limited in temporal bandwidth. As a result, deprived of a significant portion of spatiotemporal information, PAI is in constant need of better image reconstruction, processing, and analysis methods.

Traditional methods for PACT image reconstruction include Fourier domain method,^{13,14} filtered back-projection,¹⁵ delay and sum,² time reversal,^{16,17} and model-based method.^{18,19} For details about these methods, readers are referred to the review papers in Refs. 20 and 21. To recover an image with high fidelity, all existing image reconstruction methods demand wide coverage (preferably 4π solid angle), dense spatial sampling, and broadband temporal response. Such stringent requirements impose tremendous challenges to the US detection hardware, and no imaging system has met all these requirements to date, due to various technical and economic constraints. The resultant image quality reduction and poor quantification have become major roadblocks as PAI is being pushed into the clinics. Traditional image reconstruction methods have shown limited performance in recovering the lost image information under non-ideal detection. One major motivation for introducing deep learning (DL) into PAI is to improve image quality. Moreover, DL can provide a viable means to accelerate the speed of computation.

1.3 Quantitative PAI

The goal of quantitative PAI (QPAI) is to image concentrations of chromophores in tissue, thereby important physiological information can be inferred.²² For example, blood oxygen saturation is a biomarker of tumor malignancy.²³ The distinct spectral features of oxy- and deoxyhemoglobin in the near-infrared allow measurement of their concentration ratio by spectral unmixing, enabling the quantification of oxygen saturation.²⁴ In PAI, the initial pressure is a product of the local absorption coefficient, the local Gruneisen coefficient, and the local light fluence,²⁵ i.e.,

$$p_o(x, \lambda) = \Gamma(x)\phi(x, \lambda)\mu_a(x, \lambda), \quad (1)$$

where x and λ denote the spatial position and the illumination wavelength, respectively, $\Gamma(x)$ is the Gruneisen coefficient, which is a thermodynamic property of tissue, $\phi(x, \lambda)$ is the wavelength-dependent light fluence, and $\mu_a(x, \lambda)$ is the optical absorption coefficient to be determined in QPAI. Since $\phi(x, \lambda)$ and $\mu_a(x, \lambda)$ are globally coupled [any local change of $\mu_a(x, \lambda)$ tends to alter $\phi(x, \lambda)$ globally], QPAI is a highly non-linear and complex optical inversion problem. Traditional forward light propagation models include the radiative transfer equation and its approximation (e.g., diffusion approximation and δ Eddington approximation) and the Monte Carlo (MC) method for light transport.²² Existing methods for conducting optical inversion in

PAI to decouple the local fluence and absorption coefficient include linearization, direct inversion, fixed-point iteration, eigenspectra decomposition, and model-based minimization.^{22,26} Nevertheless, the accuracy, robustness, and efficiency of these methods need further validation. In recent years, DL has shown great potential in solving QPAI problems.

2 Deep Learning

2.1 DL and Convolutional Neural Networks

DL is a set of methods that utilize multiple processing layers to discover intricate structures in high-dimensional data and has made great impact in computer vision,²⁷ natural language processing,²⁸ knowledge graph,²⁹ and medical image analysis.^{30,31}

In DL, the data set used to train the network determines the generalization ability and robustness of the learning model. As a result, data set construction is always a key issue in DL. Loss function (LF) is used to measure how well the network completes the task during training. For image processing, mean square error (MSE) is the earliest and most widely applied LF. Later, some metrics for the evaluation of image quality, such as structural similarity index measure (SSIM) and peak-signal-to-noise ratio (PSNR), were introduced as loss terms. Popular architectures for DL include stacked autoencoders, deep Boltzmann machines, recurrent neural networks (RNN), and convolutional neural network (CNN). Among them, CNN is the most commonly used model in computer vision and image processing and has been studied extensively in PAI.³² A typical CNN architecture contains subsequent layers of convolution, pooling, activation, and classification (full connection). The convolution layer generates the feature graph by convolving the kernel with the input. Activation functions determine the output of some layers (e.g., convolution layer and full connection) by introducing non-linearity. Common activation functions are sigmoid, tanh, rectified linear unit (ReLU), and ReLU's variants such as Leaky. ReLU is one of the most commonly used activation functions because of its strong non-linearity and the ease of gradient calculation. The classification layer is generally a fully connected layer, which is used to connect the feature map and the output. Normalization layers can be added between the




Fig. 1 The architecture of U-Net.³⁴ Each blue box corresponds to a multi-channel feature map. White boxes represent copied feature maps. Gray arrows indicate straight-connected path on the same level.

convolutional layers and the activation functions to speed up the training process and reduce the susceptibility to network initialization. It applies a transformation that maintains the mean activation close to zero and the activation standard deviation close to one. The most commonly used normalization method is batch normalization.

So far, many architectures of CNN have been proposed. Lee et al.³³ summarized those that were widely used in medical image processing. One popular architecture is U-Net, proposed by Long and Shelhamer.³⁴ In PAI, many published works were based on this architecture or its variants.^{35–37} The basic U-Net architecture is shown in Fig. 1. The architecture consists of a contracting path to capture context and a symmetric expanding path that enables precise localization. This allows the network to effectively extract image features at different levels with outstanding performance for end-to-end tasks. The copy and crop paths (skip connections) make the network more robust and can achieve better performance with fewer data.

2.2 Tools

2.2.1 Popular DL tools

With the rapid hardware and software development, DL is being rapidly implemented in various fields. Popular open-source frameworks include: Caffe, Tensorflow, Theano, MXNet, and Torch7/PyTorch/PyTorch2 and Caffe2.^{38,39} Many of these frameworks provide multiple interfaces, such as C/C++, MATLAB, and Python. In addition, several packages provide advanced libraries written on top of these frameworks, such as Keras. These tools enabled the widespread applications of DL in various fields.

2.2.2 Tools for building data set

Since PAI has not been widely applied in the clinics, there are currently insufficient data sets for deep neural network training. Consequently, common remedies to this problem include using simulated data for proof-of-concept verification, or training the network before conducting transfer learning with real experimental data.

The PA effect involves a light scattering process followed by an US propagation process. To simulate US propagation, the k-space pseudo-spectral method can be used.⁴⁰ It combines the calculation of spatial derivatives (such as the Fourier collocation method) with a temporal propagator expressed in the spatial frequency domain or k-space. k-wave is an open source acoustics simulation toolbox for MATLAB and C++ developed by Treeby and Cox.⁴¹ It provides great flexibility and functionality in PA simulation and can be used for time reversal-based image reconstruction.⁴² Alternatively, one can solve the acoustic propagation equation using second order finite-difference time-domain (FDTD) method or finite-element method,^{43,44} to this end, COMSOL can be used.⁴⁵

To simulate the light scattering process or to perform model-based spectral unmixing in QPAI, the MC method or radiative transfer equation can be implemented as the forward model (see Sec. 1.3). The MC method is considered to be the gold standard.⁴⁶ It is also compatible with parallel processing to reduce computational time. mcxyz is a GPU-accelerated MC simulation tool to calculate light transport in optically heterogeneous tissues.⁴⁷ MCXLAB is a MATLAB package that implements the MC model. Alternatively, the radiative transfer equation or its approximation can be implemented as the forward model. For example, assuming near-diffused propagation of light dramatically reduces the need for computing power, thus fluence can be calculated numerically by FDTD or finite-element method.⁴⁶ Nirfast is an open source software that allows users to easily model near-infrared light transport in tissue.⁴⁸ COMSOL is a multi-physics simulation tool based on finite-element analysis, which can also be used to calculate light fluence distribution.⁴⁵

In PAI, currently reported training data sets have been listed in Table 1. Most of them contain MRI and x-ray CT images that can be transformed into PA images by simulation, whereas the last one provides real PA images. Detailed information about the construction of numerical phantoms for PAI can be found in Ref. 61. DL is also used for segmenting different tissue types when unlabeled data from other imaging modalities are applied for simulating PA raw data.⁶²

Table 1 Data sets commonly used in DL-based PAI

Data set	Descriptions
Mammography image database from LAPIMO EESC/USP	The database consisted of around 1400 screening mammography images from around 320 patients. ⁴⁹
DRIVE dataset	The database was used for comparative study of vascular segmentation in retinal images. It consisted of 40 photographs, 7 of which showed signs of mild early diabetic retinopathy. ⁵⁰
Optical and acoustic breast database (OA-breast)	The database includes a collection of numerical breast phantoms generated from clinical magnetic resonance angiography data collected from Washington University in St. Louis School of Medicine. ⁵¹
Digital mouse	The database includes a 3D whole body mouse atlas from coregistered high-quality PET x-ray CT and cryosection data of a normal nude male mouse. ^{52,53}
Shepp–Logan phantom	The Shepp–Logan phantom is a standard test image created by Larry Shepp and Benjamin F. Logan for their 1974 paper “The Fourier Reconstruction of a Head Section.” ⁵⁴
3D volume of CBA mouse brain vasculature	The database includes a high-resolution volumetric and vasculature atlas on CBA mouse brain based on a combination of magnetic resonance imaging and x-ray CT. ⁵⁵
ELCAP public lung image database	The database consists of an image set of 50 low-dose whole-lung CT scans. The CT scans were obtained in a single-breath hold with a 1.25-mm slice thickness. The locations of nodules detected by the radiologist are also provided. ⁵⁶
Big data from CT scanning	CT scans of several cadavers are provided. The data are collected at Massachusetts General Hospital at multiple different radiation dose levels for different x-ray spectra and with representative reconstruction techniques. ⁵⁷
Tumor phantom in mouse brain	The database is based on segmentation of a micro-CT scan of a mouse brain into gray mater, vasculature, and dura mater. An artificial cancer tissue was created by a stochastic growth process. https://github.com/asHauptmann/3DPAT_DGD/tree/master/phantomData .
VICTRE project	A series of toolkits include breast Phantom, breastCompress, and breastCrop. Using breast Phantom, the digital breast with varying patient characteristics (breast shape, glandularity and density, and size) can be generated. https://github.com/DIDSR/VICTRE .
CBIS-DDSM (curated breast imaging subset of DDSM)	The data set contains 2620 scanned film mammography studies including normal, benign, and malignant cases with verified pathology information. ^{58,59,60}
Mouse PACT ³⁷	The database consists of six athymic nude-Fox1nu mice (Harlan Laboratories) <i>in vivo</i> PA images. Each mouse was scanned over 50 mm in 0.5 mm steps, with a total of 100 cross-sectional images covering the full torso from the shoulders to the lower abdomen. https://github.com/ndavoudi/sparse_artefact_unet/blob/master/dataset/getdata.sh .

In addition, to improve preclinical image quality and accelerate PAI’s clinical translations, the international photoacoustic standardization consortium is pushing forward an open-access platform for standardized PA reference data.⁶³

3 Applications of DL in PAI

Currently, applying DL in PAI has been extensively studied, and there are already several review papers. In the review by Yang et al.,⁶⁴ they used a schematic diagram to describe the relationship between PAI tasks and the network architectures. Several open sources for implementing DL in

PAI were also listed. Andreas and Ben⁶⁵ introduced the basic principles of DL and the PA image reconstruction methods. They trained and tested the DL models with different architectures on the same data set and demonstrated and compared their performance. Gröhl et al. summarized and provided detailed statistics of the existing works in Ref. 66. Key insights for each type of work were given from the authors' perspective. This review aims at linking the technical difficulties in PAI and the DL methods that are suitable to address them. We hope to help readers understand how to build data sets and design suitable networks for their specific tasks.

The applications of DL techniques in PAI can be classified into three types of tasks: image understanding, PA initial pressure reconstruction, and quantitative PA imaging. Tasks of the first type involve using DL for PA image classification and segmentation. It also involves image registration in various anatomical structures and tissues. This type of work is covered in Sec. 3.1. Tasks of the second type focus on improving image quality under non-ideal detection conditions, where common non-ideal conditions include: limited bandwidth, sparse spatial sampling, and limited-view detection. In Sec. 3.2, we first talk about image reconstruction under a single-non-ideal detection condition. Then how DL can be used to alleviate image degradation associated with non-uniform speed of sound (SoS) will be introduced. Next, the complex scenario when multiple non-ideal conditions coexist will be discussed, such situations are more relevant to real applications. In this case, DL can be used to reconstruct PA images with better fidelity, irrespective of the specific type of the information deficiency. Finally, we introduce how DL can be used in conjunction with economical and portable PAI systems, such as single-channel data acquisition (DAQ) and light-emitting diode (LED)-based systems. For the last type of tasks, DL is used to improve the quantification accuracy and calculation speeds in QPAI. These works will be discussed in Sec. 3.3, where we first introduce how DL helps the calculation of sO_2 , chromophore concentration, and fluence spectra, then we show that DL can be used to select regions, in which QPAI calculation is reliable (which is in fact a segmentation task).

Figure 2 summarizes the structure of the above classification, which also highlights the organization of this review. Most of the references cited in this review were searched in Google Scholar with keywords “DL” and “PA”, “DL” and “optoacoustic,” “neural network” and “PA,” “neural network” and “optoacoustic,” “machine learning” and “PA,” “machine learning” and “optoacoustic”, and we tried not to miss important works by checking all relevant precursors of the original paper pool. So this review should cover DL-PAI works published before November 2020.




Fig. 2 Diagram showing the applications of DL in PAI, and the structure of this review follows this diagram.

3.1 Image Understanding

Image classification and segmentation are basic tasks in image understanding. Using DL to do auxiliary diagnosis and segmentation of lesion regions on PA images appeared in 2012.⁶⁷ Unlike traditional image segmentation, segmentation of PA images can be seen as a special reconstruction task and achieved a great result.⁶⁸ Table 2 shows the PA image understanding tasks and the networks applied. Here NN denotes “neural network” and “simple NN/CNN” refers to a NN/CNN without a complex or specific architecture, such as squeeze-and-excitation (SE)-blocks, residual layer, or dilated convolution kernel; “complex NN/CNN” refers to a NN/CNN with complex or specific architecture but without a well-known name. U-Net and its variants are called U-Net. The above naming rules apply to Tables 3 and 4 as well.

Auxiliary diagnosis is one of the most common applications of DL. As early as 2012, Alqasemi et al.⁶⁷ utilized a neural network as a classifier to facilitate ovarian cancer diagnosis on US and PACT images. In this work, features for classification were extracted from PA images by traditional methods (Fourier transform, image statistics, and composite image filters). Its performance was not as good as support vector machine (SVM). In 2016, Rajanna et al.⁶⁹ used more features from the time and frequency domains of PAM images and applied three commonly used activation functions to classify prostate cancer into three classes (malignant, benign, and normal), achieving 95.04% accuracy on average. In these works, the authors used only a simple NN as the classifier, rather than extracting features by using convolution layers or other deeper network architectures. Consequently, the NN did not exhibit superiority over traditional methods.

In 2018, Zhang et al.⁷⁰ applied AlexNet and GoogLeNet to classify breast cancer on stimulated PACT data. They achieved accuracies of 87.69% and 91.18% using AlexNet and GoogLeNet, respectively. In comparison, the accuracy of SVM was only 82.14%. Data augmentation and transfer learning were used to tackle data deficiency. They resized the images to fixed sizes and took steps to amplify the data set on preprocessing, including scaling, cropping, and random changing of the pixel value and contrast. The authors used pretrained AlexNet and GoogLeNet, and only modified and retrained the fully connected layer for these classification tasks.

Jnawali et al.⁷¹ used Inception-Resnet-V2 to detect thyroid cancer based on multi-spectral PA images and also applied transfer learning to train the model. They used a special device constituted by an acoustic lens and a rotary 1D detector array to perform C-scans. They used images at three wavelengths (760, 800, and 850 nm) as the input and achieved promising results. The area under the curve (AUC)¹³³ was 0.73, 0.81, and 0.88 for cancer, benign nodules, and normal tissue, respectively, on *ex vivo* data. Then the authors stacked twenty-one 2D PA image groups taken at five different wavelengths to form a three-dimensional (3D) PA image cube, then used a seven-layer 3D CNN and an eleven-layer 3D CNN for cancer diagnosis.⁷³ During training, the authors introduced a class-weight parameter to balance the distributions of positive and negative samples in the training data.¹³⁴ The best 3D model (11-layer 3D CNN) could detect cancer with an AUC of 0.96 and the AUC of the 2D model was 0.72, which showed that 3D features could provide more information and that deeper models seemed to have stronger feature extraction capabilities. Based on this device, Dhengre et al.⁷⁵ used several simple CNNs with the same architecture, or a combination of the CNNs with SVM or random forest (RF), to classify malignant, normal, and benign prostate hyperplasia prostate tissues (three binary classification tasks).

Table 2 Network architectures used in PA image understanding.

General task	Specific task	Network architecture
Image understanding	Classification	Simple NN; ^{67,69} AlexNet; ⁷⁰ GoogLeNet; ⁷⁰ Resnet; ^{71,72} Simple CNN; ^{73,74} simple CNN combined with traditional classifier; ^{74,75} and ResNet18 ⁷⁶
	Segmentation	U-Net ^{77,78} and simple NN applied in iterative method ⁶⁸
	Motion correction	Simple CNN ⁷⁹

The input of the network was a 105-timepoint A-line sequence in the region of interest. Involved models were: simple CNN, SVM, RF, SVM with features extracted by CNN (CNN_SVM), and RF with features extracted by CNN (CNN_RF). CNN_RF produced the highest sensitivity in most cases (highest to 0.993).

For mesoscopy imaging, Moustakidis et al.⁷⁴ used traditional machine learning methods including ensemble learning methods and DL methods to identify skin layers in PA tomograms. These images were gotten by raster-scan optoacoustic mesoscopy (RSOM).¹³⁵ In addition to applying simple CNN, they also used modified Alexnet and Resnet whose fully connected layer and classification layer were replaced by principal component analysis (PCA) and RF to classify 3D image directly. The simple CNN provided a classification accuracy of ~85%, which was the best result in DL models. Nitkunanantharajah et al. used ResNet18¹³⁶ to diagnose systemic sclerosis by identifying microvascular changes at the finger nailfold.⁷⁶ The microvascular was also imaged by RSOM and segmented manually. The network's input was the 2D images generated by frequency band equalization¹³⁷ and maximum intensity projection. Transfer learning was applied to train the model, which ultimately achieved an AUC of 0.897, a sensitivity of 0.783, and a specificity of 0.895.

DL also played an important role in PA image segmentation, where U-Net is the most successful model. Chlis et al.⁷⁷ applied an adapted sparse U-Net to perform vascular segmentation on clinical multi-spectral PA images. They used a 1×1 2D convolution layer to transform an image with 20 spectral channels (wavelengths from 700 to 970 nm in 10 nm steps) to a single-channel image. The model was trained on *in vivo* data acquired using a handheld PA/US system. The ground truth was established based on the consensus between two clinical experts. The performance of the sparse U-Net was similar to the standard U-Net whose size was 30 times bigger. They found that sparse U-Net was faster thus was more suitable for clinical application. Berkan et al.⁷⁸ also showed the U-Net's performance on segmentation of the mouse boundary in an (optoacoustic US) OPUS system. PA signal was detected by a 270-deg ring array and images were reconstructed by BP. They used manually segmented images as the ground truth. The dice coefficients on the cross-sectional images of the brain, liver, and kidney were 0.98, 0.96, and 0.97, respectively.

Boink et al.⁶⁸ proposed a learned primal-dual (L-PD) model for simultaneous PA reconstruction and segmentation. Primal-dual algorithms are popular in tomography reconstruction.^{138,139} In their work, Boink et al. used a CNN-based model instead of the primal-dual hybrid gradient to learn the best update in each iteration during model-based image reconstruction. It consists of two networks: one for calculating the previous dual update and the other for calculating the primal update, which corresponds to the initial pressure. Compared to convex segmentation¹⁴⁰ and U-Net, the L-PD method yielded the best results. There were two major innovations: (1) the data set was augmented by changing the number of detectors. (2) Deep-learning-based reconstruction was used to perform image segmentation. In other words, the authors reconstructed a segmentation map.

Motion artifacts and pixel dislocation are almost inevitable in optical resolution PAM of *in vivo* targets. These are attributable to the breathing motion and heartbeat of the animal, or to the positioning errors of the motor. Motion correction can improve the performance of image understanding. Chen et al.⁷⁹ proposed a simple CNN with three convolutional layers to do motion correction. They employed it to process *in vivo* rat brain vessel images contaminated by motion artifacts. They found that the model's performance was significantly improved using a larger kernel size, at the expense of longer processing time. Compared with several existing algorithms that are not based on DL,^{141–144} the proposed model demonstrated the best performance.

3.2 Reconstruction of Initial Pressure Images

Traditional PA image reconstruction methods were introduced in Sec. 1. In brief, ideally a full detection solid angle of 4π , sufficient spatial sampling frequency, infinite detector bandwidth, and known distribution of the SoS, are the conditions for the traditional reconstruction methods to work properly. But these conditions are rarely satisfied in real applications—US transducers are non-ideal, and tissue properties can be complex and unknown—consequently, available




Fig. 3 PA reconstruction quality is influenced by different non-ideal conditions: (a) gold standard; (b) good-quality reconstruction (DAS); reconstruction with reduced quality due to (c) limited bandwidth; (d) sparse sampling; (e) limited view; and (f) wrong SoS.

information is inadequate for high-quality image reconstructions. The effects of information deficiency on the image quality are shown in Fig. 3, where the target was obtained from the DRIVE data set and simulated by k-wave.¹⁴⁵ The applied reconstruction method is DAS. In some cases, signal quality is sacrificed for various purposes such as system cost. For example, LED are used to replace bulky lasers for system miniaturization.

DL methods were introduced to compensate for the above-mentioned signal deficiency. These methods can be classified into five types: (1) preprocessing of channel data: the NN is used to process the raw data, which is subsequently fed into a traditional reconstruction program to produce an image. (2) Postprocessing of reconstructed images: the network is applied to boost the quality of the images reconstructed by traditional methods. (3) Direct reconstruction of the initial pressure: the network directly outputs the initial pressure image when channel data are fed as the input. (4) Combining image reconstruction and postprocessing: the network improves image quality based upon both low-quality images and raw channel data. (5) Embedding DL in traditional reconstruction frameworks: use DL to perform certain calculations in traditional reconstruction methods (for example, to calculate the update or regularization term in model-based methods). Table 3 summarizes the network architecture for each kind of task and each type of model.

The DL-based methods are often compared against traditional methods in terms of SSIM, PSNR, Pearson correlation coefficient (PCC), mean absolute error (MAE), MSE, normalized 2D cross-correlation (NCC), signal-to-noise ratio (SNR), contrast-to-noise ratio (CNR), and edge preserving index. These quantities were also added into the LF for better training.

3.2.1 Single-non-ideal detection condition

Signal detections in PAI are often limited in bandwidth, sampling density, and angular coverage. For example, although linear arrays are commonly used in clinical settings, they suffer from all these limitations thus could only produce images with relatively low quality.¹⁴⁶ In this section, we review how DL is applied to improve the image quality of PAI, where the role of DL is to compensate for the information loss due to one of the following: limited bandwidth, sparse sampling, and limited-view angle.

Table 3 Network architectures used in PA image reconstruction

	Non-ideal condition	Pre-processing	Postprocessing	Direct reconstruction	Combined reconstruction	Embedded in traditional reconstruction
Single-non-ideal detection	Limited bandwidth	Simple NN ⁸⁰		U-Net ^{81,82}		
	Sparse sampling		U-Net ⁸³⁻⁸⁵ and FD-Unet ^{86,87} Simple CNN ⁸⁹ and complex CNN ⁹⁰			Simple CNN ⁸⁸
	Limited view		U-Net ^{91,92} and VGG ⁹²	U-Net ⁹¹ and complex CNN ⁹³	Multiple branches autoencoder ³⁶	
Non-uniform SoS			Faster R-CNN ⁹⁴⁻⁹⁷ U-Net ^{99,100}	Simple CNN ⁹⁸		
Multiple non-ideal conditions	U-net ¹⁰¹	U-Net; ^{32,37,102} S-Net; ³² WGAN (based on U-Net); ¹⁰³ and simple CNN ¹⁰⁴		U-Net ¹⁰⁵ and multiple branches autoencoder ¹⁰⁶	Ki-GAN (based on multiple branches autoencoder) ¹⁰⁷	U-Net; ¹⁰⁸⁻¹¹⁰ FD-Unet; ¹¹¹ simple CNN; ^{68,112} multiple branches autoencoder; ¹¹³ and RNN ¹¹⁴
Use of DL in economical channel and portable PAI devices	Single-channel DAQ			Autoencoder (based on LSTM) ¹¹⁵		
	LED-PAI		Complex CNN ¹¹⁶ and CNN&LSTM ¹¹⁷	U-Net ¹¹⁸⁻¹²²		

Limited bandwidth. Transducer bandwidth affects both the axial and lateral resolutions.¹⁴⁷ Commercially available US detectors for medical imaging are typically narrowband, in contrast, PA signals are broadband, spanning from several tens of kilohertz to beyond a hundred megahertz.²⁰ Although capacitive micromachined ultrasonic transducers and optical transducers may have larger bandwidth, they have not been widely adopted for PAI yet.¹⁴⁸ A reduced bandwidth inevitably results in the loss of important image features.

In 2017, Gutte et al. used a five-layer fully connected deep NN to broaden the bandwidth of the channel data. The input and output of the network were both channel data (sinograms).⁸⁰ The authors trained the network on numerical breast phantoms and tested its performance by simulated data (blood vessel network, Derenzo phantom, etc.) and real phantom data (horse hair phantom and ink tube phantom). Compared with least-squares deconvolution in terms of PCC, CNR, and SNR, the DL method almost tripled these measures in all the tests. Plus, it was almost 1.63 times faster than the least-squares deconvolution. In addition to deconvolving the frequency response, DL was also used to synthesize a broader frequency response given multi-band signals. Lan et al.⁸¹ used ConvU-Net, which was a modified U-Net with a kernel size of (20×3) and a stride of (20×1) as skipped connections, to reconstruct PA images from sinograms interleaved with signals individually acquired at three different center frequencies: 2.25, 5, and 7.5 MHz. It is a direct reconstruction model: the input of this network was the raw signal from 120 channels, containing three 40-channel subgroups that were distributed on a circle corresponding to the three center frequencies and the output was the reconstructed images. The authors trained and tested the network on numerical phantoms containing segmented vessels from fundus oculi CT

imaging.¹⁴⁵ ConvU-Net showed better image reconstruction quality than TR, DAS, and common U-Net in terms of SNR, PSNR, SSIM, and relative error. Lan et al.⁸² then used two U-Net structures connected end-to-end to reconstruct images from the three frequencies data. The model outperformed ConvU-Net in a series of tests.

Sparse sampling. The minimum spatial density to arrange detector elements is determined by the Nyquist sampling criterion. The problem in which the actual detector density is lower than the minimum required density is called sparse sampling. In many applications, spatial sampling is sparse, due to various reasons such as system cost and imaging speed. Visually, sparse sampling introduces streak artifacts and may jeopardize image resolution. Postprocessing models are often applied to address such problems.

In 2017, Antholzer et al.⁸³ used a U-Net to process PA images that had been reconstructed by FBP based on sparse sampling data. The data were detected by 30 transducers evenly arranged around a circle. The authors trained their network on simulated data where Gaussian noise was added. Compared with the results obtained using FBP and the model-based method on simulation data, the DL method yielded the best quality. In addition, postprocessing by U-Net reconstruction required only 20 ms (FBP: 15 ms and U-Net: 5 ms) compared with 25 s for the model-based method. Guan et al. proposed a modified U-Net, termed FD-UNet, in which each layer is connected with every other layer. It was also a postprocessing model to remove streak artifacts. The original PA images were reconstructed by time reversal from sparse data.⁸⁶ The training and testing data sets were generated from three different digital phantoms (circles, Shepp–Logan, and experimentally acquired micro-CT images of mouse brain vasculature) on different levels of sampling sparsity. FD-UNet performed obviously better than U-Net on PSNR and SSIM. When applied to the mouse brain vasculature data, FD-UNet outperformed U-Net by recovering more details. After transfer learning was applied, the networks' performance was improved after fine-tuning with small and well-matched training data set. In 2020, Farnia et al.⁸⁴ used a U-Net to process a PA image reconstructed using time reversal. Deng et al.⁸⁵ used SE-UNet, a modified U-Net with added SE-blocks¹⁴⁹ in skip connections, to remove artifacts in PA images reconstructed by BP. It got better results than BP and Sta-UNet.⁹¹

For PAM, DiSpirito et al.⁸⁷ used FD-UNet to reconstruct images of *in vivo* mouse brain microvasculature blurred by under sampling. The fully sampled images were used as the ground truth and the images down-sampled at a ratio of 5:1 in one scan direction were employed as the input. Compared with U-Net, ResU-Net, and ResICL U-Net,¹⁵⁰ FD-UNet had the best performance in PSNR and MSE and was the second best in SSIM. FD-UNet also showed good robustness, it did not require high image contrast, and performed stably well for images with different down-sampling ratios. Zhou et al.⁹⁰ proposed a CNN-based model to improve the quality of PAM images suffering sparse sampling. Sixteen residual blocks and eight SE blocks were used for feature extraction after the first convolution layer of CNN. They claimed that the residual blocks performed well in super-resolution tasks and the SE blocks helped convergence. Compared with Bicubic interpolation and EDSR (another CNN-based method),¹⁵¹ the new method showed the best PSNR and SSIM on PAM images of leaf veins and *in vivo* data (blood vessels of the mouse eyes and ears).

Except U-Net, Awasthi et al.⁸⁹ used a seven-layer network to achieve super-resolution and denoising on the channel data. The training dataset was simulated by the CHASE,¹⁵² DRIVE,¹⁴⁵ and STARE¹⁵³ databases. Their simulation involved 100 detectors, and Gaussian noise (SNR to 20/40/60 dB) was added. They then spatially down-sampled the data by a factor of 2, resulting in a reduced channel number of 50. Finally, the down-sampled data was interpolated to recover a total channel number of 100 as the input. The output of the network was the residual between the interpolation result and the initial signal detected by all the detectors. On average, the network reduced RMSE by 41.70% while increased PSNR by 6.93 dB on simulated data (numerical blood vessel and Derenzo phantoms) and *in vivo* data (rat brain).

Image reconstruction incorporating DL and model-based methods have been also explored. Antholzer et al.⁸⁸ developed a simple NN consisting of three convolutional layers to calculate the regularization term in model-based image reconstruction, termed network Tikhonov (NETT). They built a training set of input/output pairs, in which the inputs were images

reconstructed by FBP from under sampling data and the outputs were the differences between the predicted signal and the ground truth to train the regularization. When the input was the ground truth, the output was zero. Compared with FBP, a compressed sensing method with l_1 -minimization,¹⁵⁴ a model-based method with H^1 -regularization, and U-Net (used as a post-processing method), NETT produced the best results with noise-free data and satisfactory results with noisy data.

In this section, we show how DL can remove artifacts generated by sparse sampling. Similar approaches were reported by Schwab et al.¹⁵⁵ and Guan et al.¹¹¹ Since these works deal with situations beyond merely sparse sampling, they will be discussed in Sec. 3.2.3.

Limited-view detection. Incomplete angular coverage can distort image features by partially losing information and generating artifacts. The problem is frequently encountered in PAI and is termed “limited view.” Starting from 2018, DL has been extensively studied to solve the limited-view problem. Because in the missing cone signal is completely lost, tackling limited view is more of a super-resolution than a deconvolution problem.

In 2018, Waibel et al. investigated both postprocessing and direct reconstruction to solve the limited-view problem.⁹¹ They used a U-Net to do postprocessing, and a modified U-Net with an additional convolutional layer that resized the information to the target resolution in skip connections, to reconstruct PA image directly. In their simulation, the PA signals generated by circular and elliptical targets were detected by a linear array. The median relative estimation error improved from 98% (DAS) to 10% (postprocessing) and 14% (direct reconstruction). The direct reconstruction method was faster than postprocessing because it did not need the DAS step. Deng et al. proposed a postprocessing method, in which they built an *in vivo* data set and used two DL models to improve *in vivo* image quality. They utilized the PA signal acquired by a ring array to generate the data set.⁹² The PA images reconstructed from a full ring were used as the ground truth, and images reconstructed from a partial ring (1/4 ring, 90-deg coverage) were used as the images with limited-view artifacts. (A similar approach was demonstrated by Davoudi et al.³⁷ and will be discussed in Sec. 3.2.5.) They used U-Net to postprocess the images reconstructed by DAS. They also mapped the full-view images, all taken at the liver region of the mouse, into a multi-dimensional space (20 in their work) and calculated the bases and their weights by PCA. VGG was then used to establish the mapping between the projection weights of the full-view and the limited-view images. These models all produced satisfactory results.

For direct reconstruction, Anas et al.⁹³ proposed a model on the basis of CNN. They used dilated and large convolution kernels at higher layers to obtain global information to prevent resolution loss. They performed simulations (using circular targets) to train, validate, and test the network. This model could obtain a nearly 8.3 dB increase of PSNR. Lan et al. proposed a hybrid autoencoder model, called Y-Net. It consists of two intersecting encoder paths that got information from reconstructed images or raw data to solve the limited-view problem in linear array PAI systems. The authors compared the performance of Y-Net with that of time reversal, DAS, and U-Net (postprocessing) in terms of SSIM, PSNR, and SNR on simulated data and experimental data (chicken breast tissue inserted with two pencil leads and a *in vivo* human palm), Y-Net got the best results. Moreover, Y-Net achieved a better “transfer” ability than U-Net because it referred to information in the raw data.

There are works that addressed both the sparse sampling and limited-view issues as will be introduced in Sec. 3.2.3.

3.2.2 Non-uniform speed of sound

The SoS in biological tissues is generally non-uniform, with unknown spatial distributions. In addition, a SoS mismatch often exists between the biological tissue and its surrounding medium. In all cases, if instead a constant SoS is assumed during the image reconstruction process, artifacts will be generated (such as feature splits). Moreover, a large SoS discontinuity (e.g., between muscle and bone) is bound to create acoustic reflections. The reflected signals can mix with the PA signal to introduce reflection artifacts. Recent studies have shown the great potential of DL in solving the acoustic heterogeneity problem in PAI. Researchers first treated non-uniform SoS as a target identification problem, addressing whether an image feature is real or fake.

Other researchers used DL to reconstruct images directly or to perform postprocessing on PA images originally generated using traditional methods.

In 2017, Reiter and Bell⁹⁸ trained an eight-layer simple CNN to predict realistic images (point targets) in heterogeneous media directly from raw data. The model achieved satisfying results (mean axial and lateral point location errors of 0.28 and 0.37 mm, respectively) at various sound speeds (1440 to 1640 m/s), target locations (5 to 25 mm), and absorber sizes (1 to 5 mm) using simulated data. This work predicted the point locations successfully but failed to identify whether the wave front was from the actual target or the reflection of the target. Allman et al. implemented a series of faster R-CNN models based on VGG, Resnet-50, and Resnet-101 to decide whether the identified features were actual targets or reflection artifacts in PA images reconstructed by DAS.⁹⁴⁻⁹⁷ Faster R-CNN is an object detection framework based on deep convolutional networks. It can provide bounding boxes and confidence scores for target detection. In their works, faster R-CNN output the type of the objects (i.e., source or artifact), along with their locations in the form of the coordinates of the bounding box and a confidence score between 0 and 1 for each object. The authors trained their models on simulated data and tested them using phantom data, *in vivo* data (an optical fiber in a pig vessel) and *ex vivo* data (a needle tip inserted into tissue). PA signals were detected using a linear array. Faster R-CNN was able to accurately identify and locate the targets. All three DL models produced satisfactory results on phantom data and *ex vivo* data. For *in vivo* data, the residual network architectures correctly classified 83.3% (Resnet-50) and 88.8% (Resnet-101) of the sources. The authors suspected that a deeper network (i.e., residual network) has a greater capacity to learn higher-level features.

Other than point targets, DL was also studied to deal with acoustic heterogeneity for general targets and these are all postprocessing models. In 2019, Shan et al.⁹⁹ used a modified U-Net for postprocessing images after the first iteration of a model-based algorithm. The input of the U-Net was the feature map extracted from the PA images by a four-layer CNN, whereas the output of the U-Net was processed by a network consisting of four deconvolution layers as the final reconstructed image. They added image structure information (SSIM) in the LF. The authors trained the network and tested its performance using data generated based on cadaver CT scans. In their simulations, they randomly picked a circular region where SoS was set to be high and introduced a sinusoidal change of SoS in the background tissue. And 0%, 10%, 20%, 30%, and 40% noise was added to the PA signals. The DL method produced the best results compared to that of TR and other popular iterative algorithms, such as averaged time reversal (ATR),¹⁵⁶ adjoint ATR,¹⁵⁷ and Landweber iteration.¹⁵⁸ In addition, without doing iterations, the DL method was the fastest among all compared algorithms. To compensate for the SoS heterogeneity, Jeon and Kim¹⁰⁰ proposed an autoencoder model whose architecture was similar to U-Net. The data were also contaminated by noise and beamformed with various SoS to generate the multi-channelled input data (eight channels with eight speeds). The ground truth image was obtained by reconstructing the noise-free data with the correct SoS. The model was able to remove image artifacts associated with acoustic heterogeneity, even though it had never been trained with the correct SoS map on *in vivo* data (forearm). In addition, the model was also effective in reducing side lobes and noise.

3.2.3 Multiple non-ideal conditions

In Secs. 3.2.1 and 3.2.2, we have shown that DL can be used in applications where a single-non-ideal detection condition dominates. In real cases, often times the combination of several such conditions may conspire to reduce the image quality. We have shown that certain DL networks, such as U-Net and its variants, were successfully applied to deal with individual factors including sparse sampling, limited view, and acoustic heterogeneity. Consequently, such networks are expected to be applicable when several of these factors simultaneously take effect. Here we introduce DL methods that work in conditions where multiple non-ideal factors co-exist. These techniques are more promising for practical uses.

In this section, we will first introduce the preprocessing and postprocessing models. Second, direct reconstruction models will be discussed. Third, we will introduce the combined models that use raw data and reconstructed image as inputs. Finally, we will introduce how DL can be

used in conjunction with traditional reconstruction methods, for example, DL networks can be used to calculate the sum in DAS or compute the update in model-based methods.

For preprocessing models, Awasthi et al.¹⁰¹ proposed a modified U-Net for super-resolution and bandwidth extension. Named U-Net (hybrid) and directly working on the channel data, the proposed network used exponential linear units (ELUs) as the activation function in the final layers and used ReLUs in other layers.⁶¹ The input of the network was the signal detected by 100 bandwidth-limited detectors, and then the data were interpolated to 200 channels employing the nearest neighbor method. The ground truth data were generated with 200 full-bandwidth detectors. They divided the channel data into small subsections, the size of each subsection was 64×64 . The network only processed one such subsection at one time. Compared with SRCNN, U-Nets (implementing only ReLU or ELU), direct interpolation, and automated wavelet denoising,¹⁵⁹ the proposed network got the highest SNR for phantom data (horse hair) and *in vivo* data (rat brain).

In an attempt to apply DL to the postprocessing of PA images, Antholzer et al.³² used a U-Net and an S-Net (simple network consisting of three layers) to enhance PA images originally reconstructed by FBP in limited-view and sparse-sampling conditions. In the numerical study, 24 sensors were located on a non-closed curve (less than a semicircle). Both networks were successful in removing most of the artifacts. The more complex U-Net produced better results but also took a longer time to train and process the images. Davoudi et al.³⁷ also employed U-Net to perform postprocessing on sparse sampling and limited-view images. They obtained a full-view image from a full-view tomographic scanner with 512 transducers in a circle as the ground truth. Sparse-sampling and limited-view data were gotten by down sampling. The training data set was constructed using images of 6 mice with 100 different cross sections from each. Their model produced satisfying results on *in vivo* data, as shown in Fig. 4. The authors also made their data set available online.

Godefroy et al.¹⁰² introduced the Bayesian machine learning framework into PA image post-processing Bayesian machine learning framework into PA. They applied MC dropout as a Bayesian approximation in U-Net.¹⁶⁰ So the model was able to generate 20 different outputs for the same input. The input was the images reconstructed by DAS, and the prediction image was the mean of the different outputs. In their experiment, PA images taken by a linear array were used as the input, and the photographs taken by a CMOS camera were used as the ground truth. Test results showed that the model effectively improved NCC and sSSIM and was more robust than networks without dropout.

Vu et al.¹⁰³ used Wasserstein generative adversarial network with gradient penalty (WGAN-GP) to do postprocessing on PA images to remove artifacts due to limited-view and limited-bandwidth detection. The generator in WGAN-GP took the form of a U-Net, and the discriminator of this model was a common CNN. The discriminator was trained to identify the ground truth from the generator's output. The generator's input was the PA image reconstructed by TR. They used a complex LF to train the network. It was the sum of three factors: (1) the distance between the discriminator's output and the true distribution. The Wasserstein distance was employed here to take advantage of its capability in describing morphological characteristics. (2) The GP, which could avoid gradient explosion or disappearance. (3) The MSE loss, which helped preserve the information in the reconstructed images. In simulation, phantom, and *in vivo* experiments, WGAN-GP all showed slightly better performance on SSIM, PSNR, and in recovering low-contrast structures than U-Net.

Zhang et al.¹⁰⁴ developed a nine-layer CNN to solve the under-sampling and limited-view problems on PA images reconstructed by UBP. The PA signal was detected by a three-quarter ring transducer array with 128 elements. The network was trained on simulated 3D vascular patterns. They also trained and tested a compressed sensing model and in all numerical, phantom, and *in vivo* experiments, the CNN got better results.

For direct reconstruction, in 2020, Feng et al.¹⁰⁵ used a Res-UNet network to reconstruct PA images. In their work, signals were detected by 128 transducers covering an angle of 270 deg around the imaged object. They added a residual connection on the layer of a U-Net and a convolution layer was introduced as a copy-and-crop path in the network. Generalization of the ResU-Net network was achieved by imaging six types of synthetic phantoms, including discs, breads, spiders, lines, logos, and natural pictures. They compared the network's performance




Fig. 4 The results of Davoudi et al.'s method. (a) Reconstructed image with under sampled (32 projections) data versus its artifact-free counterpart obtained with the trained network. TP, true positive; FP, false positive; TN, true negative; and FN, false negative. (b) Another example for under sampled data with 128 projections. (c) Top: image reconstructed with 60-deg angular coverage and its respective amplitude spectrum. Bottom: output image of the network and its corresponding amplitude spectrum. (d) Top: image reconstructed with 135-deg angular coverage and its respective amplitude spectrum. Bottom: output image of the network and its corresponding amplitude spectrum.

with U-Net on digital and physical phantoms. Res-UNet got the best PC and PSNR on digital phantom and the images with the least artifacts on physical phantom. Inspired by the fact that FBP takes the time derivative of the signal as input, Tong et al.¹⁰⁶ proposed a feature projection network (FPnet) to reconstruct PA images based on limited-view and sparsely sampled data. Its architecture resembled that of Y-net with the two input branches fed with the channel data and the time derivative of the channel data, whereas the output was the initial pressure image. Then they used U-Net to do postprocessing on the image. The outputs of FPnet and U-Net were trained on simulated and *in vivo* data to match the ground truth. The hybrid model showed better results compared with: (1) FPnet only, (2) postprocessing (by U-Net) images reconstructed by FBP, and (3) other traditional methods (FBP and model-based).

Lan et al. proposed a knowledge infusion generative adversarial network (Ki-GAN). As a combined network, Ki-GAN had two branches, which got information from reconstructed images and the raw data.¹⁰⁷ Compared with Y-net, Ki-GAN had a more complex architecture.

A knowledge embedding branch (KEB, which is a CNN with a three-layer inception block architecture) was used to extract information in the generator. Information from the two branches were merged by a four-layer CNN to reconstruct the final image. The architectures of Ki-GAN and KEB are shown in Figs. 5(a) and 5(b). The discriminator penalized the texture at the scale of patches.¹⁶¹ The model implemented an information-rich LF in the form of the linear combination of several MSE loss terms to improve feature extraction. Compared to Ki-GAN with DAS, U-Net (including direct reconstruction and postprocessing), and the iteration method (10 steps), Ki-GAN yielded the best results in terms of SSIM, PSNR, and SNR. The results of the *in vivo* experiment are shown in Fig. 5(c). Meanwhile, Ki-GAN was much faster than the iteration method (0.025 versus 331.51 s).

In traditional PA image reconstruction, certain computation tasks can be accomplished using DL for stronger fitting capability and faster speed. For example, DL can be applied to learn the parameters of traditional reconstruction methods. Dynamic aperture length (DAL) correction is one solution to the non-ideal detection problem.^{162,163} In 2018, Schwab et al. proposed a framework named DALnet, in which an image was first reconstructed by UBP with DAL correction, and then processed by a CNN network (U-Net with a residual connection) to address the limited-view and under-sampling problem.¹⁰⁸ The weights of the DAL correction and the U-Net were jointly trained. The output of the model was compared with those of UBP and a model-based




Fig. 5 The architecture of Lan et al.'s model. (a) The overall architecture of Ki-GAN; KEB represents convolutional layers; DAS: delay and sum reconstruction. (b) The detailed architecture of KEB. (c) PA images of rat thigh reconstructed by iterative algorithm with 10 iterations (column 1), U-Net (column 2), and Ki-GAN (column 3).

method in terms of total variation, and DALnet was superior in both image quality and computing speed. After this work, Schwab et al. also proposed a similar network to learn the weights in UBP to improve the PAT image quality under limited-view and sparse-sampling conditions.¹⁵⁵ The network only had two layers. The first layer received raw data as input and carried out temporal filtering without any trainable weights, and the second layer performed back projection with adjustable weights. The following conditions were considered: limited view, sparse sampling, and limited-view plus sparse sampling. The network-assisted UBP reduced the relative errors of conventional UBP from 0.2002, 0.3461, and 0.3545 to 0.0912, 0.1806, and 0.1649, in the three conditions, respectively.

Other approaches of combining FBP/DAS with DL have been developed as well. Under the non-ideal detection conditions, DL has been used in FBP/DAS to replace the sum operation to restore some of the lost information. Guan et al.¹¹¹ proposed a DL model called pixel-wise DL (Pixel-DL) for image reconstruction under limited-view and sparse sampling conditions. In Pixel-DL, they back-projected the interpolated channel data to the image grid using a constant SoS. Then the projected data group consisted of multiple channels were put into FD-UNet for image reconstruction. The authors compared the performance of Pixel-DL to those of other PAT image reconstruction methods, including TR, iterative reconstruction, post-DL (using FD-UNet to process an image reconstructed by TR), and modified Direct-DL (using FD-UNet to reconstruct PA image from raw data). Pixel-DL generated comparative results with the iterative method, and both were better than the rest. In 2019, Kim et al.¹⁰⁹ employed the same concept and proposed a model using U-Net to reconstruct images in a linear array setting. The authors reformatted the detected PA signals into multi-channel 3D (tensorial) data using prior knowledge about the acoustic propagation delays and utilized a U-Net model to produce reconstructed images. Compared with DAS, DMAS,^{164,165} model-based, and U-Net (for postprocessing), the new model produced the best results when applied to simulated data. For the phantom and *in vivo* data, the model provided superior contrast and resolution.

The model-based methods can be used to reconstruct images in the non-ideal detection conditions. In some cases, high-quality images can be produced and used as the ground truth.¹⁰⁸ However, implementation of model-based reconstructions is often iterative, making them time-consuming; to accelerate their speeds, DL models can be used.

Hauptmann et al.¹¹² proposed a deep NN that learned the entire iteration process of a model-based reconstruction algorithm for limited view and sparse sampling. The method was called deep gradient descent (DGD). The current result and the gradient of the LF were used to calculate the output of the next iteration by using a five-layer CNN. The network used a greedy approach, in which every network output was used to match the ground truth. The planar array used in the study inherently had a limited view, and sparse sampling was performed by a 16× random selection of transducer elements. When applied to simulated data, after two iterations DGD was able to match the image quality of the traditional iterative method after 50 iterations. Transfer learning was used to advance the training for the *in vivo* data. They used fully sampled images reconstructed by the model-based method as the gold standard. The DGD after retraining produced satisfactory results. In addition, the iterative methods seemed to be more robust than U-Net, which was more sensitive to data variations. Later, Hauptmann et al.¹¹⁰ proposed a modified DGD, called fast-forward PAT, which employed a multi-scale network, like a shallow U-Net, to update the results. According to the test, with 4× subsampled data, FF-PAI with five iterations produced comparable results to the model-based method with 20 iterations. Compared with the conventional model-based method, FF-PAI reduced the computation time by a factor of 32. Boink et al.⁶⁸ proposed an L-PD model to perform image reconstruction and segmentation simultaneously using the same network, and their method has been introduced in Sec. 3.1. The result of image reconstruction and segmentation was generated in the first and second channel, respectively. According to the authors, L-PD produced better results than FBP, model-based method with TV, and postprocessing with U-Net, when noise and sparse sampling were concerned.

Shan et al.¹¹³ developed a CNN-based model to reconstruct the SoS and initial pressure jointly using the framework of the model-based method. The updates of initial pressure and SoS were computed based on three types of input: (1) the initial pressure distribution, (2) the negative gradient of the L2 distance between the reconstructed image and the ground truth, and

(3) the current SoS distribution. Feature extraction, feature fusion, and image reconstruction were all processed by the CNN. The model produced satisfactory results, where the MAE for SoS and the initial pressure were, respectively, $(0.85 \pm 0.33) \times 10^{-2}$ Pa and 1.24 ± 0.52 m/s (mean \pm std) after four iterations on simulation data.

Yang et al.¹¹⁴ proposed a recurrent inference machine (RIM) based on an RNN to calculate the update in each iteration step. The reconstruction results and the gradient of the objective function (optimization equation) were used as inputs for the next iteration step. Gated recurrent unit was used in this network.¹⁶⁶ This unit has an update gate and a reset gate to avoid information loss. The first iteration input was calculated by multiplying the detected signal by the adjoint of the acoustic transmission matrix. The authors compared their results with those obtained using the DGD algorithm with five iterations, and RIM produced slightly better result. Compared with U-Net, RIM reduced the number of parameters by almost fourteen times.

3.2.4 Use of DL in economical and portable PAI devices

DL was applied in some PAI systems to help achieve device miniaturization, cost reduction, and image quality improvement. We will first introduce how DL helps to reconstruct images based on single-channel DAQ. Then we will show how DL is used to enhance image quality in situations where the light intensity is low (e.g., when LED are used), and the role of DL can be seen as performing segmentation between target and noise. Again, U-Net and its variants are the most popular architectures used.

In 2019, Lan et al.¹¹⁵ developed a PACT system to provide real-time imaging. The imaging system consisted of 120 sensing elements connected to a DAQ unit with only a single channel. In their system, the 120-channel signals were overlaid directly into 4-channel signals before being fed into a delay-line module, allowing the signals to be temporally separable and combined into a single channel. Subsequently, the final output (single channel) could be separated into four channels, from which the authors used a DL network to reconstruct the PA images. The architecture of the network was also an autoencoder where the encoder was LSTM and the decoder was CNN. They were connected by fully connected layers. Compared to DAS with 120 channels, the proposed model could perform image reconstruction faster (28 versus 159 ms) with satisfactory result on phantom data (four black balls in agarose gel). This demonstrated the potential of the single-channel DAQ for high-speed low-cost operation.

Pulsed laser diodes (PLD) and LED are stable, affordable, and compact alternative light sources for PAI. However, their output energy is low, typically in the range of nJ/pulse to μ J/pulse.¹¹⁸ The low energy leads to low signal intensity and thus reduced image quality. A favorable property of PLD/LED is that they have high repetition frequency, typically on the order of kHz and above. A common method for denoising is averaging, at the expense of temporal resolution. Yet, even with averaging the image quality of the PLD/LED-based systems is suboptimum compared to their bulk-laser-based counterparts.

In 2018, Anas et al. proposed a model that consisted of series of dense convolutional layers¹⁶⁷ to improve the quality of a 2D LED-based PA image.¹¹⁶ In addition to the final output, it also has two outputs in the middle layers that were all fed into LF to train model. A number of raw data frames were averaged and reconstructed by DAS as input. The model could increase frame rate by 6 times compared to the conventional averaging approach, with a mean PSNR of 34.5 dB and a mean SSIM of 0.86. Then Anas et al.¹¹⁷ combined CNN and ConvLSTM to improve the quality of these images. Its architecture is shown in Fig. 6(a). The inputs of the network were a sequence of PA images. The model used CNN to extract spatial features, whereas ConvLSTM¹⁶⁸ was employed to exploit the temporal correlation among the images. The input is a sequence of averaged PA images. In a phantom experiment (phantom: a wire and a tube filled with gold nanoparticles), the network increased the frame rate by 8.1 times compared to the CNN-only method.

Kerrick et al.¹¹⁹ also proposed an encoder–decoder with atrous convolution (called as Nyquist convolution) on input layer to predict the location of circular chromophore targets in tissue mimicking a strong scattering background. It can achieve micron level accuracy on simulated data. For *in vivo* data, the network significantly improved the image quality.




Fig. 6 The architecture of (a) Anas et al.'s network and (b) Johnstonbaugh et al.'s network. Details of the residual convolution module and the upsampling module are also provided.

In 2019, Johnstonbaugh et al.¹²⁰ also proposed an encoder–decoder model for the direct reconstruction of point targets under low illumination intensity. The model was based on U-Net whose architecture is shown in Fig. 6(b). In each of the last two layers, a high field-of-view convolution module (5×5) was applied in between the encoder and decoder. The down sampling designs (Nyquist convolution) reduced the network size and sped up the calculations. The model successfully imaged a deep vessel target with an accurate estimation of its position on AcousticX system, while the corresponding image reconstructed by FBP was noisy and difficult to interpret. The model also produced good multi-target localization results.

For postprocessing tasks, U-Net was typically used. Hariri et al. used a modified U-Net, termed multi-level wavelet-CNN, in which the pooling operations were replaced by discrete wavelet transform (DWT), and pooling operations were replaced by DWT, and the upsampling operations were replaced by inverse wavelet transform, to enhance PA image quality in a low SNR setting.¹¹⁸ They used the PA images taken at a fluence of 17 mJ/pulse as the ground truth and then reduced the laser fluence down to 0.95 and 0.25 mJ/pulse to train the network. In *in vivo* experiment, mice injected with various concentrations of methylene blue (MB) were imaged, and the CNR improvement factor was 1.55, 1.76, 1.62, and 1.48 for a dye concentration of 0.05, 0.1, 1.0, and 5.0 mM, respectively. Similarly, Singh et al. used a U-Net network to enhance the image quality of LED-based PAI.¹²¹ The targets were small tubes filled with MB or indocyanine green (ICG) in a water bath. PA images were obtained using one LED system and two OPO systems. The network was trained with images obtained by the two OPO systems and tested with images acquired by the LED-based system. The DL network improved the SNR by $\sim 30\%$. Manwar et al.¹²² also reported a similar idea and used U-Net to enhance the SNR of deep structures in brain tissue. They got B-scan images from an *ex vivo* sheep brain by a linear array at 20 and 100 mJ as the input and the label, respectively. They also evaluated several LFs including MAE, MSE, SSIM, multi-scale SSIM (MS-SSIM), and some combinations of these factors (MS-SSIM + MSE and MS-SSIM + L1). The network can enhance image quantity effectively, and the networks with combined LF got better performance. To improve SNR, some novel network structures can be used, such as denoising convolutional neural network.¹⁶⁹

3.3 Quantitative Photoacoustic Imaging

With known spectral profiles of the major tissue chromophores, the goal of QPAI is to image the concentrations of the molecular targets through a process known as the optical inversion. Because of wavelength-dependent light absorption and scattering, the fluence spectrum in deep tissue is unknown, thus the absorption spectrum cannot be directly inferred from the PA spectrum. Traditional QPAI strategies often rely on overly ideal assumptions, such as piecewise constant optical properties, *a priori* knowledge of scattering coefficients, and homogeneous (and known) background optical properties.^{22,35} Among the developed methods, model-based iterative

Table 4 Network architectures used in QPAI

Task categories	Network architecture
Calculate sO_2 , chromophore concentration, or fluence spectra	U-Net, ^{35,123-129} Autoencoder, ¹³⁰ Simple NN, ¹³¹ and LSTM&CNN applied in traditional methods ¹³²
Generate mask for trusted zones	U-Net ¹²⁷⁻¹²⁹

optimization can provide relatively accurate solutions but are time-consuming and sensitive to quantification errors in the PA images.^{170,171} Diffuse optical tomography can facilitate fluence distribution estimation but tends to increase system complexity and cost.¹⁷² In recent years, DL methods were applied to solve the QPAI problem with great success. The main reason why NNs are suitable for QPAI is that they are good at solving non-linear problems. Autoencoder with the end-to-end architecture is suitable for QPAI, in which the input is the PA images acquired at different wavelengths, whereas the output is the concentration map of the target molecules. In these tasks, U-Net and its variants as well as fully connected networks were commonly applied. Most of the works involved estimating the concentration maps for the whole tissue or manually segmented regions of interest. However, the prediction results were poor in areas with weak optical absorption. It would be beneficial to extract trusted regions with sufficiently high SNR for quantitative analysis. The DL-assisted QPAI tasks and applied networks are shown in Table 4.

In 2017, Kirchner et al.¹⁷³ used an ML method based on random forest regression to perform fluence estimation and tested its performance using simulated data. The method relied only on the local PA signal near the target and did not make use of either the global information or any feature extraction schemes such as deep NNs. In 2018, Cai et al.³⁵ applied an end-to-end NN for the simultaneous imaging of oxygen saturation (sO_2) and concentration of ICG using multi-wavelength PA images as input, taking advantage of DL's strong ability to represent complex mapping. Their network, shown in Fig. 7(a), was a modified U-Net with embedded residual structures. According to their simulation, the mean estimation errors of sO_2 and ICG concentration were 0.76% and 3.26% for the circular phantom, and 0.51% and 7.51% for the digital mouse. The imaging result of the digital mouse is shown in Figs. 7(b) and 7(c). This model exhibited certain immunity to noise with relative error as low as 1.43% (20 dB SNR) and 16.8% (10 dB SNR). Hoffer-Hawlik et al.¹²³ developed absO2luteU-Net, which used a new activation function with ELU instead of the ReLU in U-Net to calculate sO_2 based on dual-color PA image (700 and 900 nm). The model produced an RMSE and SO-RMSE of 4.49% and 18.4%, respectively, compared to 75.5% and 64.8% by linear unmixing. AbsO2luteU-Net




Fig. 7 (a) The ResU-Net architecture implemented by Cai et al. (b) Simultaneously reconstructed sO_2 map (gray) and ICG concentration (color). (c) Absolute sO_2 error (gray) and relative ICG concentration error (color). Scale bars: 5 mm.

exhibited surprisingly low-noise sensitivity, the pixel-wise RMSE remained $<6\%$ when the SNR varied from 0 to 20 dB. Chen et al.¹²⁴ used a U-Net with the leaky ReLU activation function to calculate optical absorption. The input of network is the reconstructed images obtained at a single wavelength, and the output is the optical absorption map. The LF used was the summation of MSE and the TV regularization term. The relative error of the outputs in different absorption backgrounds was $<10\%$.

In addition to the simple U-Net, Yang et al. presented a combination model, called deep residual and recurrent NN, to estimate sO_2 from the initial pressure images taken at two wavelengths.¹²⁵ The networks used complex layers including convolution branch and residual branch in U-Net. According to simulation, the error generated by the network was as low as 1.43% compared with 62.39% by linear unmixing. Because only two wavelengths were used, the estimation process took only 18.4 ms. In 2019, Yang et al.¹²⁶ proposed a complex autoencoder called EDA-net, which was intended specifically for QPAI to achieve accurate quantification of sO_2 from PA images acquired at 21 wavelengths (700 to 800 nm in 5-nm steps). EDA-net had an encoder path to extract information, a decoder path to accept feature images and calculate sO_2 , and several aggregation nodes to mix information. They were built by convolutional layers. Each layer of the encoder path was connected to each layer of the decoder path by aggregation nodes. Compared with linear unmixing, ResU-Net, U-Net++,¹⁷⁴ and EDA-net were superior. The authors also found that when the number of wavelengths was increased to more than nine, the mean error did not decrease significantly. In addition to this automatic encoder structure (including U-Net and its variants), Gröhl et al.¹³¹ used a nine-layer fully connected NN to reconstruct sO_2 from initial pressure images with 26 wavelengths (700 to 950 nm, step size 10 nm) in 2019. The model operated in a pixel-wise manner, meaning that local information, rather than global information, was employed. For *in vivo* data of porcine brain and human forearm, the network's output of 90% and 98% was close to the expected arterial blood oxygenation values of healthy subjects compared with linear unmixing (68% and 80%, respectively). This was the first attempt to apply DL on *in vivo* QPAI data, and the results were superior to those obtained using linear spectral unmixing.

Durairaj et al.¹³⁰ proposed an unsupervised learning approach for PA spectral unmixing. Their model included two networks: an initialization network and an unmixing network. The weights of the initialization network were used as the initial weights for the unmixing network. The initialization network was a pixel-wise model. The number of input and output nodes was equal to the number of wavelengths (six in their case), and the number of hidden nodes was equal to the number of targets (three in their case, which were HbO_2 , ICG, and Hb). The LF was the L1 norm of the difference between the input and the output. The unmixing network had more channels and used the whole set of multi-spectral images as input. The numerical phantom had three molecular targets (HbO_2 , ICG, and Hb). In terms of estimation accuracy, the model was similar to linear unmixing, but it did not need prior knowledge about the absorption spectra.

Eigenspectra multi-spectral optoacoustic tomography is a model to calculate light fluence in deep tissue.¹⁷⁵ Olefir et al.¹³² combined the eigenspectra concept and DL and named the new method as DL-eMSOT. They used LSTM and CNN to calculate the weights of four spectral bases for predicting the eigenfluence. The eigenfluence of every pixel was generated by interpolation and sO_2 was calculated by linear unmixing. In the *in vivo* experiment, they inserted a tube filled with porcine blood of known oxygenation (0% or 100%) into the animal. For most cases, DL-eMSOT outperformed eMSOT. Furthermore, the calculation speed of DL-eMSOT was 60 times faster.

For generating mask of trusted regions, Gröhl et al.¹²⁷ presented a framework in which estimation confidences were used to increase the quantification accuracy. They used a CNR map corresponding to the initial pressure image and a U-Net network to create a joint mask. The input of the network was multi-spectral PA images and the output was a relative error map of optical absorption. It only selected the overlap regions between the two maps for evaluation to increase the quantification accuracy. To prove that their method was effective, the authors applied it to three different QPAI methods: naïve fluence correction (using a simple MC to calculate the light fluence and simulated the initial pressure), fluence correction (using U-Net to calculate the light fluence and the initial pressure), and direct absorption estimation (using U-Net to calculate the

absorption coefficient directly). The results showed an accuracy increase of nearly 80% when applying a 50% confidence threshold in the direct absorption estimation model, which also produced the best results among the three models. Luke et al. presented O-net, in which two U-Nets shared the same input and were combined to estimate the vascular sO_2 and segment the blood vessels. The input of the network was dual-wavelengths PA data.¹²⁸ In their results, sO_2 was estimated only in the blood vessels (as segmented by the ground-truth oxygen maps) when the network was trained. The model exhibited certain immunity to noise and produced absolute errors of 5.1% and 13.0% when the SNR was 25 and 5 dB, respectively. Bench et al. used two simple U-Nets for the segmentation and sO_2 estimation of 3D PA images.¹⁷⁶ The network was trained on a data set generated numerically using 3D vessel models acquired from CT scans of human lung vessels. In the simulation, a skin model consisting of the epidermis, dermis, and hypodermis layers with different optical properties was applied. They used the segmentation results to generate a mask to calculate the mean sO_2 and found the mean difference between the ground truth and the output to be 0.3%, with a standard deviation of 6.3%.

4 Discussion

4.1 Network Architecture and LF

As discussed earlier, the information loss due to deficient signal detection has been a major roadblock for the development of PAI.

According to the published results, we find that: (1) U-Net is almost a panacea that has been applied in most of the tasks; (2) the application of DL in the traditional reconstruction frameworks usually generates better results than other approaches; and (3) the better results are often accompanied by increased network complexity, such as more branches,^{36,126} more connections,^{86,87} and more complex layers or blocks.^{107,149}

Is there a more fundamental design principle? Can we design networks based on physics and generate interpretable structures that correspond to the first-principal process? Since interpretability is associated with the trustworthiness, expandability, and sustainable development of the DL approaches, complementing the data-driven approaches with explainable ingredients or top-down designs has been a major endeavor for current and future research.¹⁷⁷

Supplementing the missing information is also one of the major aims of network design. Combining DL with traditional reconstruction methods is a good choice, which has yielded good results in all types of task.^{68,109,110,132} Among them, applying DL in iterative model-based methods to calculate the update or the regularization term has been the most successful, probably because it can best leverage the prior information about the PA physics and the imaged object. In addition, the iterative method is more robust than other models (including preprocessing model, postprocessing model, and direct reconstruction model).¹¹²

People often use MSE as a common LF, but more items (such as SSIM, PSNR, and PC^{99,103}) can be added for various purposes such as better information extraction and faster convergence. Notably, GAN is an emerging network^{103,107} whose LF contains a network that is being simultaneously trained. The GAN network has been shown to recover fine details better.

Real-time data processing and display are important for clinical applications, and currently most PAI systems are based on personal computers or FPGAs. Limited by the available computing resources, people have to pay close attention to the computational complexity and the overall size when designing the network. The exploratory works in the CV field, such as SqueezeNet,¹⁷⁸ MobileNet,¹⁷⁹ and ShuffleNet,¹⁸⁰ may be inspiring for the development of future DL-assisted PAI systems with small size and high speed.

4.2 Data Set

The availability of high-quality data set is of paramount importance to the success of the DL methods. A prominent example is the development of CNNs, which were made possible by ImageNet.¹⁸¹ However, PAI is an emerging technology lacking high-quality data sets. The following remedies were used in the PAI community to alleviate the problem.

- *Using simulated data.* It is a common method to build data sets, which is covered in Sec. 2.2.2.
- *Data augmentation.* It can be used to generate more data from the existing data. Common methods include random pixel shifting, rotation, cropping, warping, vertical and horizontal flipping, and adding noise.^{70,72}
- *Transfer learning.* It is a commonly used technique, in which the network is pretrained on public/simulated data before being retrained on a more relevant, high-quality data set with limited size.^{70,112}
- *Network downsizing.* It can be used to reduce the input size. For example, when the input is cropped from 256×256 to 64×64 , the size of the data set is equivalently increased.^{100,101} In addition, it can reduce the number of network parameters so that the network can be trained with fewer data.
- *Unsupervised model.* It is an alternative option where no ground truth is needed.¹³⁰

Despite the above-mentioned solutions, large, high-quality PA data sets are bound to facilitate the development of DL-PAI. On the one hand, a high-quality data set makes network development more efficient, reliable, and flexible. On the other hand, there is no effective way to compare the performance of different models. This is to a large extent because the systems and methods used by different research groups are vastly diverse, and intermural standards for the imaging data and evaluation metrics do not exist. Most importantly, building a high-quality and sizable database tailored for PAI is necessary. Fortunately, some organizations have begun building such data sets.⁶³

Currently, building data sets and applying them in real-world applications remain challenging. Since DL models are unexplainable, one is unsure whether a model is solely dependent on the physics of image formation, or it is also affected by specific image features. One must be aware that once the network parameters are coupled with image features, it can hallucinate based on what it has learnt, so one must confirm the result by comparing it with the ground truth. Because the image features of PA are unique, it is difficult to obtain a gold standard image using other modalities such as MRI, and without such ground-truth knowledge, how can we know that the network is working properly? This is especially true for QPAI, since PAI is currently the only imaging modality for sO_2 and chromophore concentration measurement (with high spatial resolution in great depth). Thus due to the lack of a reliable QPAI computation method, it is currently impossible to obtain the ground truth in live animals or humans. Developing realistic phantoms for PAI and QPAI is thus an important future research direction.

In real applications, data balancing is also noteworthy. This means that the training set distributions will affect the predicted results.^{182,183} It is a general rule of thumb that the proportion of each data type is consistent with that in the actual application. If the condition is unsatisfied, applying weight parameters is a useful method to balance the data.⁶⁹

4.3 Reliability of Results

As we have briefly discussed in Sec. 4.2, in what condition, and to what extent can we trust an image reconstructed by DL methods? For example, we have shown that DL can be used to improve the image quality under limited-view detection. Since in the missing cone all information is completely lost, what the NN does is to fill in the lost information as in super-resolution, rather than to amplify weak frequency components as in deconvolution. Since a physical mechanism for super-resolution is lacking (such as introducing non-linearity into the image formation process), the true reason for the recovery of the missing frequency components may be that the network recognizes the image features during training. This means that a network trained with images acquired at the liver region would produce unrealistic features in the kidney by hallucinating. We expect that even for the imaging of the same organ (e.g., the breast), the network may produce fake results if the target being imaged has features that are new to the network. The situation will become worse when a network trained on one system is translated to another, especially when the two systems have different types of probes and/or imaging targets. Thus the study and verification of the generalizability of the DL networks are important.

Tradition methods also perform well on promoting the PA image quality under non-ideal detection.^{184,185} Generally speaking, they are currently more reliable than the DL methods. Comparison between and integration of the traditional methods and the DL methods are important future research directions.

4.4 Applications

Compared with the traditional modalities, such as CT, US, and MRI, PAI is an emerging imaging modality, in which most clinicians do not have much experience. It is easier for doctors to accept PAI if DL can be used to extract key image features or relevant physiological parameters for auxiliary diagnosis.

For some clinical applications, information loss is inevitable. For hand-held probes, it is difficult for common ultrasonic medical probes to receive wide-band PA signals (e.g., 0.1 to 10 MHz). Shallow features are prone to be affected by the sparse sampling, whereas the deep features are more subject to limited view. In these situations, DL is an effective means to compensate for the lost information. DL can also facilitate the development of low-cost and portable equipment, which usually have reduced image quality. In such applications, building the data set is comparatively easier and DL will play an important role. For example, one can get high-quality images using powerful lasers as the ground truth.

In addition to providing information about vascular morphology,¹⁸⁶ a unique capability provided by PA is the imaging of chromophore concentrations and sO₂ in tissue. Due to the lack of a closed-form solution in these problems, data-based approaches (not limited to DL) have achieved promising results.¹⁷⁵ We expect DL to play an increasingly important role in QPAI.

Specific devices incorporating DL can achieve good imaging results unachievable otherwise.¹¹⁵ It is believed that, combined with DL, novel PAI devices can be designed to meet the needs of various applications. In particular, as 3D PAI is becoming increasingly popular,^{65,72,104,129} using DL for 3D reconstruction, generating 3D images from 2D data, or processing 3D images are important future directions.

4.5 Conclusion

Despite the above challenges and difficulties, we envision that DL will continue to make great impact for PA imaging. The unparalleled capability of DL in information extraction, fusion, and high-speed processing is bound to bring PAI new vigor and opportunities. This review is thus not a conclusion, but rather the herald of the exciting era of DL-based PA imaging.

Disclosures

The authors have no relevant financial interests in this article and no potential conflicts of interest to disclose.

Acknowledgments

This research was funded by the National Natural Science Foundation of China (Nos. 61735016 and 61971265). The authors would like to thank Youwei Bao and Xiangxiu Zhang for assistance with figure copyright application.

References

1. J. Laufer, "Photoacoustic imaging: principles and applications," in *Quantification of Biophysical Parameters in Medical Imaging*, I. Sack and T. Schaeffter, Eds., pp. 303–324, Springer International Publishing, Cham (2018).
2. J. Xia, J. Yao, and L. V. Wang, "Photoacoustic tomography: principles and advances," *Electromagn. Waves* **147**, 1–22 (2014).

3. I. Steinberg et al., “Photoacoustic clinical imaging,” *Photoacoustics* **14**, 77–98 (2019).
4. A. Rosencwaig and P. Griffiths, “Photoacoustic and photoacoustic spectroscopy,” *Phys. Today* **34**, 64 (1981).
5. J. Yao and L. Wang, “Photoacoustic microscopy,” *Laser Photonics Rev.* **7**, 758–778 (2013).
6. J. Yao et al., “Noninvasive photoacoustic computed tomography of mouse brain metabolism in vivo,” *NeuroImage* **64**, 257–266 (2012).
7. A. A. Oraevsky et al., “Clinical optoacoustic imaging combined with ultrasound for coregistered functional and anatomical mapping of breast tumors,” *Photoacoustics* **12**, 30–45 (2018).
8. Y. Zeng et al., “Photoacoustic and ultrasonic coimage with a linear transducer array,” *Opt. Lett.* **29**, 1760–1762 (2004).
9. L. Li et al., “Single-impulse panoramic photoacoustic computed tomography of small-animal whole-body dynamics at high spatiotemporal resolution,” *Nat. Biomed. Eng.* **1**, 0071 (2017).
10. G. Diot et al., “Multispectral optoacoustic tomography (MSOT) of human breast cancer,” *Clin. Cancer Res.* **23**, 6912–6922 (2017).
11. M. Toi et al., “Visualization of tumor-related blood vessels in human breast by photoacoustic imaging system with a hemispherical detector array,” *Sci. Rep.* **7**, 41970 (2017).
12. K. Fukutani et al., “Characterization of photoacoustic tomography system with dual illumination,” *Proc. SPIE* **7899**, 78992J (2011).
13. Y. Xu, D. Feng, and L. V. Wang, “Exact frequency-domain reconstruction for thermoacoustic tomography—I: planar geometry,” *IEEE Trans. Med. Imaging* **21**(7), 823–828 (2002).
14. L. Kunyansky, “Fast reconstruction algorithms for the thermoacoustic tomography in certain domains with cylindrical or spherical symmetries,” *Inverse Prob. Imaging* **6**, 111–131 (2012).
15. M. Xu and L. V. Wang, “Universal back-projection algorithm for photoacoustic computed tomography,” *Phys. Rev. E* **71**(1 Pt. 2), 016706 (2005).
16. Y. Xu and L. Wang, “Time reversal and its application to tomography with diffracting sources,” *Phys. Rev. Lett.* **92**, 033902 (2004).
17. Y. Hristova, P. Kuchment, and L. Nguyen, “Reconstruction and time reversal in thermoacoustic tomography in acoustically homogeneous and inhomogeneous media,” *Inverse Prob.* **24**, 055006 (2008).
18. G. Paltauf et al., “Iterative reconstruction algorithm for optoacoustic imaging,” *J. Acoust. Soc. Am.* **112**, 1536–1544 (2002).
19. A. Rosenthal, D. Razansky, and V. Ntziachristos, “Fast semi-analytical model-based acoustic inversion for quantitative optoacoustic tomography,” *IEEE Trans. Med. Imaging* **29**, 1275–1285 (2010).
20. C. Lutzweiler and D. Razansky, “Optoacoustic imaging and tomography: reconstruction approaches and outstanding challenges in image performance and quantification,” *Sensors* **13**, 7345–7384 (2013).
21. R. Amir, N. Vasilis, and R. Daniel, “Acoustic inversion in optoacoustic tomography: a review,” *Curr. Med. Imaging* **9**(4), 318–336 (2013).
22. B. Cox et al., “Quantitative spectroscopic photoacoustic imaging: a review,” *J. Biomed. Opt.* **17**, 061202 (2012).
23. P. Vaupel, F. Kallinowski, and P. Okunieff, “Blood flow, oxygen and nutrient supply, and metabolic microenvironment of human tumors: a review,” *Cancer Res.* **49**(23), 6449–6465 (1989).
24. M. Li et al., “Simultaneous molecular and hypoxia imaging of brain tumors in vivo using spectroscopic photoacoustic tomography,” *Proc. IEEE* **96**(3), 481–489 (2008).
25. L. Wang, H.-I. Wu, and B. Masters, “Biomedical optics: principles and imaging,” *J. Biomed. Opt.* **13**, 049902 (2008).
26. M. Li, Y. Tang, and J. Yao, “Photoacoustic tomography of blood oxygenation: a mini review,” *Photoacoustics* **10**, 65–73 (2018).
27. A. Voulodimos et al., “Deep learning for computer vision: a brief review,” *Comput. Intell. Neurosci.* **2018**, 1–13 (2018).

28. T. Young et al., “Recent trends in deep learning based natural language processing [review article],” *IEEE Comput. Intell. Mag.* **13**(3), 55–75 (2018).
29. K. S. Valluru and J. K. Willmann, “Clinical photoacoustic imaging of cancer,” *Ultrasonography* **35**(4), 267–280 (2016).
30. H.-M. Zhang and B. Dong, “A review on deep learning in medical image reconstruction,” *J. Oper. Res. Soc. China* **8**, 311–340 (2020).
31. K. Suzuki, “Overview of deep learning in medical imaging,” *Radiol. Phys. Technol.* **10**(3), 257–273 (2017).
32. S. Antholzer et al., “Photoacoustic image reconstruction via deep learning,” *Proc. SPIE* **10494**, 104944U (2018).
33. J.-G. Lee et al., “Deep learning in medical imaging: general overview,” *Korean J. Radiol.* **18**, 570 (2017).
34. O. Ronneberger, P. Fischer, and T. Brox, “U-net: convolutional networks for biomedical image segmentation,” *Lect. Notes Comput. Sci.* **9351**, 234–241 (2015).
35. C. Cai et al., “End-to-end deep neural network for optical inversion in quantitative photoacoustic imaging,” *Opt. Lett.* **43**(12), 2752–2755 (2018).
36. H. Lan et al., “Y-net: hybrid deep learning image reconstruction for photoacoustic tomography in vivo,” *Photoacoustics* **20**, 100197 (2020).
37. N. Davoudi, X. L. Deán-Ben, and D. Razansky, “Deep learning optoacoustic tomography with sparse data,” *Nat. Mach. Intell.* **1**(10), 453–460 (2019).
38. S. Liu et al., “Deep learning in medical ultrasound analysis: a review,” *Engineering* **5**, 261–275 (2019).
39. S. Bahrapour et al., “Comparative study of deep learning software frameworks,” arXiv:1511.06435 (2015).
40. B. T. Cox and P. C. Beard, “Fast calculation of pulsed photoacoustic fields in fluids using k-space methods,” *J. Acoust. Soc. Am.* **117**, 3616–3627 (2005).
41. B. Treeby and B. Cox, “k-Wave: MATLAB toolbox for the simulation and reconstruction of photoacoustic wave fields,” *J. Biomed. Opt.* **15**(2), 021314 (2010).
42. B. Treeby, E. Zhang, and B. Cox, “Photoacoustic tomography in absorbing acoustic media using time reversal,” *Inverse Prob.* **26**, 115003–115020 (2010).
43. B. Parvitte et al., “Quantitative simulation of photoacoustic signals using finite element modelling software,” *Appl. Phys. B* **111**, 383–389 (2013).
44. Y.-L. Sheu et al., “Photoacoustic wave propagation simulations using the FDTD method with Berenger’s perfectly matched layers,” *Proc. SPIE* **6856**, 685619 (2008).
45. C. Sowmiya and A. Thittai, “Simulation of photoacoustic tomography (PAT) system in COMSOL(R) and comparison of two popular reconstruction techniques,” *Proc. SPIE* **10137**, 101371O (2017).
46. X. Zhou et al., “Evaluation of fluence correction algorithms in multispectral photoacoustic imaging,” *Photoacoustics* **19**, 100181 (2020).
47. T. Li, S. Jacques, and S. Prah, “mcxyz.c, a 3D Monte Carlo simulation of heterogeneous tissues” (2017).
48. M. A. Mastanduno and S. S. Gambhir, “Quantitative photoacoustic image reconstruction improves accuracy in deep tissue structures,” *Biomed. Opt. Express* **7**(10), 3811–3825 (2016).
49. B. R. N. Matheus and H. Schiabel, “Online mammographic images database for development and comparison of CAD schemes,” *J. Digital Imaging* **24**(3), 500–506 (2011).
50. <https://drive.grand-challenge.org/DRIVE/>.
51. Y. Lou et al., “Generation of anatomically realistic numerical phantoms for photoacoustic and ultrasonic breast imaging,” *J. Biomed. Opt.* **22**(4), 041015 (2017).
52. B. Dogdas et al., “Digimouse: a 3D whole body mouse atlas from CT and cryosection data,” *Phys. Med. Biol.* **52**, 577–587 (2007).
53. D. Stout et al., “Creating a whole body digital mouse atlas with PET, CT and cryosection images,” *Mol. Imaging Biol.* **4**(4), S27 (2002).
54. L. A. Shepp and B. F. Logan, “The Fourier reconstruction of a head section,” *IEEE Trans. Nucl. Sci.* **21**(3), 21–43 (1974).

55. A. Dorr, J. G. Sled, and N. Kabani, “Three-dimensional cerebral vasculature of the CBA mouse brain: a magnetic resonance imaging and micro computed tomography study,” *NeuroImage* **35**(4), 1409–1423 (2007).
56. VIA Research Group, Cornell University, “VIA/I-ELCAP public access research database,” 2010, <https://www.via.cornell.edu/databases/lungdb.html>.
57. Q. Yang et al., “Big Data from CT Scanning” (2015).
58. R. S. Lee et al., “A curated mammography data set for use in computer-aided detection and diagnosis research,” *Sci. Data* **4**(1), 170177 (2017).
59. M. Heath et al., “The digital database for screening mammography,” in *Proceedings of the Fifth International Workshop on Digital Mammography*, M. J. Yaffe, Ed., pp. 212–218, Medical Physics Publishing (2001).
60. M. Heath et al., “Digital database for screening mammography: 1998,” in *Digital Mammography*, pp. 457–460, Kluwer Academic Publishers; Proceedings of the Fourth International Workshop on Digital Mammography (1998).
61. Y. Lou et al., “Generation of anatomically realistic numerical phantoms for photoacoustic and ultrasonic breast imaging,” *J. Biomed. Opt.* **22**(4), 041015 (2017).
62. Y. Ma et al., “Human breast numerical model generation based on deep learning for photoacoustic imaging,” in *42nd Annu. Int. Conf. IEEE Eng. Med. and Biol. Soc.*, pp. 1919–1922 (2020).
63. S. Bohndiek et al., “International photoacoustic standardisation consortium (IPASC): overview (conference presentation),” *Proc. SPIE* **10878**, 108781N (2019).
64. C. Yang et al., “Review of deep learning for photoacoustic imaging,” *Photoacoustics* **21**, 100215 (2021).
65. H. Andreas and T. C. Ben, “Deep learning in photoacoustic tomography: current approaches and future directions,” *J. Biomed. Opt.* **25**(11), 112903 (2020).
66. J. Gröhl et al., “Deep learning for biomedical photoacoustic imaging: a review,” *Photoacoustics* **22**, 100241 (2021).
67. U. Alqasemi et al., “Recognition algorithm for assisting ovarian cancer diagnosis from coregistered ultrasound and photoacoustic images: ex vivo study,” *J. Biomed. Opt.* **17**, 126003 (2012).
68. Y. Boink, S. Manohar, and C. Brune, “A partially learned algorithm for joint photoacoustic reconstruction and segmentation,” *IEEE Trans. Med. Imaging* **39**, 129–139 (2019).
69. A. R. Rajanna et al., “Prostate cancer detection using photoacoustic imaging and deep learning,” *Electron. Imaging* **2016**, 1–6 (2016).
70. J. Zhang et al., “Photoacoustic image classification and segmentation of breast cancer: a feasibility study,” *IEEE Access* **7**, 5457–5466 (2019).
71. K. Jnawali et al., “Transfer learning for automatic cancer tissue detection using multispectral photoacoustic imaging,” *Proc. SPIE* **10950**, 109503W (2019).
72. K. Jnawali et al., “Deep 3D convolution neural network for CT brain hemorrhage classification,” *Proc. SPIE* **10575**, 105751C (2018).
73. K. Jnawali et al., “Automatic cancer tissue detection using multispectral photoacoustic imaging,” *Int. J. Comput. Assist. Radiol. Surg.* **15**(2), 309–320 (2020).
74. S. Moustakidis et al., “Fully automated identification of skin morphology in raster-scan photoacoustic mesoscopy using artificial intelligence,” *Med. Phys.* **46**(9), 4046–4056 (2019).
75. N. Dhengre et al., “Computer aided detection of prostate cancer using multiwavelength photoacoustic data with convolutional neural network,” *Biomed. Signal Process. Control* **60**, 101952 (2020).
76. S. Nitkunanantharajah et al., “Three-dimensional photoacoustic imaging of nailfold capillaries in systemic sclerosis and its potential for disease differentiation using deep learning,” *Sci. Rep.* **10**(1), 16444 (2020).
77. N. Chlis et al., “A sparse deep learning approach for automatic segmentation of human vasculature in multispectral photoacoustic tomography,” *Photoacoustics* **20**, 100203 (2020).
78. L. Berkan et al., “Efficient segmentation of multi-modal photoacoustic and ultrasound images using convolutional neural networks,” *Proc. SPIE* **11240**, 112402N (2020).
79. X. Chen, W. Qi, and L. Xi, “Deep-learning-based motion-correction algorithm in optical resolution photoacoustic microscopy,” *Visual Comput. Ind. Biomed. Art* **2**, 12 (2019).

80. S. Gutta et al., “Deep neural network-based bandwidth enhancement of photoacoustic data,” *J. Biomed. Opt.* **22**, 116001 (2017).
81. H. Lan et al., “Deep learning approach to reconstruct the photoacoustic image using multi-frequency data,” in *IEEE Int. Ultrason. Symp.*, pp. 487–489 (2019).
82. H. Lan et al., “Reconstruct the photoacoustic image based on deep learning with multi-frequency ring-shape transducer array,” in *41st Annu. Int. Conf. IEEE Eng. Med. and Biol. Soc.*, pp. 7115–7118 (2019).
83. S. Antholzer, M. Haltmeier, and J. Schwab, “Deep learning for photoacoustic tomography from sparse data,” *Inverse Prob. Sci. Eng.* **27**, 987–1005 (2019).
84. P. Farnia et al., “High-quality photoacoustic image reconstruction based on deep convolutional neural network: towards intra-operative photoacoustic imaging,” *Biomed. Phys. Eng. Express* **6**(4), 045019 (2020).
85. J. Deng et al., “Unet-based for photoacoustic imaging artifact removal,” in *Imaging and Appl. Opt. Cong.*, OSA Technical Digest, JTh2A.44 (2020).
86. S. Guan et al., “Fully dense UNet for 2D sparse photoacoustic tomography artifact removal,” *IEEE J. Biomed. Health. Inf.* **24**, 568–576 (2020).
87. A. DiSpirito et al., “Reconstructing undersampled photoacoustic microscopy images using deep learning,” *IEEE Trans. Med. Imaging* **40**(2), 562–570 (2021).
88. S. Antholzer et al., “NETT regularization for compressed sensing photoacoustic tomography,” *Proc. SPIE* **10878**, 108783B (2019).
89. N. Awasthi et al., “Sinogram super-resolution and denoising convolutional neural network (SRCN) for limited data photoacoustic tomography,” arXiv:2001.06434 (2020).
90. J. Zhou et al., “Photoacoustic microscopy with sparse data by convolutional neural networks,” *Photoacoustics* **22**, 100242 (2021).
91. D. Waibel et al., “Reconstruction of initial pressure from limited view photoacoustic images using deep learning,” *Proc. SPIE* **10494**, 104942S (2018).
92. H. Deng et al., “Machine-learning enhanced photoacoustic computed tomography in a limited view configuration,” *Proc. SPIE* **11186**, 111860J (2019).
93. E. M. A. Anas et al., “Robust photoacoustic beamforming using dense convolutional neural networks,” *Lect. Notes Comput. Sci.* **11042**, 3–11 (2018).
94. D. Allman, A. Reiter, and M. L. Bell, “A machine learning method to identify and remove reflection artifacts in photoacoustic channel data,” in *IEEE Int. Ultrason. Symp.* (2017).
95. D. Allman et al., “Deep neural networks to remove photoacoustic reflection artifacts in ex vivo and in vivo tissue,” in *IEEE Int. Ultrason. Symp.* (2018).
96. D. Allman et al., “A deep learning-based approach to identify in vivo catheter tips during photoacoustic-guided cardiac interventions,” *Proc. SPIE* **10878**, 108785E (2019).
97. D. Allman, A. Reiter, and M. A. L. Bell, “Photoacoustic source detection and reflection artifact removal enabled by deep learning,” *IEEE Trans. Med. Imaging* **37**(6), 1464–1477 (2018).
98. A. Reiter and M. L. Bell, “A machine learning approach to identifying point source locations in photoacoustic data,” *Proc. SPIE* **10064**, 100643J (2017).
99. H. Shan, G. Wang, and Y. Yang, “Accelerated correction of reflection artifacts by deep neural networks in photo-acoustic tomography,” *Appl. Sci.* **9**, 2615 (2019).
100. S. Jeon and C. Kim, “Deep learning-based speed of sound aberration correction in photoacoustic images,” *Proc. SPIE* **11240**, 112400J (2020).
101. N. Awasthi et al., “Deep neural network based sinogram super-resolution and bandwidth enhancement for limited-data photoacoustic tomography,” *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **67**, 2660–2673 (2020).
102. G. Godefroy, B. Arnal, and E. Bossy, “Compensating for visibility artefacts in photoacoustic imaging with a deep learning approach providing prediction uncertainties,” *Photoacoustics* **21**, 100218 (2021).
103. T. Vu et al., “A generative adversarial network for artifact removal in photoacoustic computed tomography with a linear-array transducer,” *Exp. Biol. Med.* **245**(7), 597–605 (2020).
104. H. Zhang et al., “A new deep learning network for mitigating limited-view and under-sampling artifacts in ring-shaped photoacoustic tomography,” *Comput. Med. Imaging Graphics* **84**, 101720 (2020).

105. J. Feng et al., “End-to-end Res-UNet based reconstruction algorithm for photoacoustic imaging,” *Biomed. Opt. Express* **11**(9), 5321–5340 (2020).
106. T. Tong et al., “Domain transform network for photoacoustic tomography from limited-view and sparsely sampled data,” *Photoacoustics* **19**, 100190 (2020).
107. H. Lan et al., “Ki-GAN: knowledge infusion generative adversarial network for photoacoustic image reconstruction in vivo,” *Lect. Notes Comput. Sci.* **11764**, 273–281 (2019).
108. J. Schwab et al., “Real-time photoacoustic projection imaging using deep learning,” arXiv:1801.06693 (2018).
109. M. Kim et al., “Deep-learning image reconstruction for real-time photoacoustic system,” *IEEE Trans. Med. Imaging* **39**, 3379–3390 (2020).
110. A. Hauptmann et al., “Approximate k-space models and deep learning for fast photoacoustic reconstruction,” *Lect. Notes Comput. Sci.* **11074**, 103–111 (2018).
111. S. Guan et al., “Limited-view and sparse photoacoustic tomography for neuroimaging with deep learning,” *Sci. Rep.* **10**(1), 8510 (2020).
112. A. Hauptmann et al., “Model based learning for accelerated, limited-view 3D photoacoustic tomography,” *IEEE Trans. Med. Imaging* **37**(6), 1382–1393 (2018).
113. H. Shan, G. Wang, and Y. Yang, “Simultaneous reconstruction of the initial pressure and sound speed in photoacoustic tomography using a deep-learning approach,” *Proc. SPIE* **11105**, 1110504 (2019).
114. C. Yang, H. Lan, and F. Gao, “Accelerated photoacoustic tomography reconstruction via recurrent inference machines,” in *41st Annu. Int. Conf. IEEE Eng. Med. and Biol. Soc.*, pp. 6371–6374 (2019).
115. H. Lan et al., “Real-time photoacoustic tomography system via single data acquisition channel,” arXiv:2001.07454 (2020).
116. E. M. A. Anas et al., “Towards a fast and safe LED-based photoacoustic imaging using deep convolutional neural network,” *Lect. Notes Comput. Sci.* **11073**, 159–167 (2018).
117. E. M. A. Anas et al., “Enabling fast and high quality LED photoacoustic imaging: a recurrent neural networks based approach,” *Biomed. Opt. Express* **9**(8), 3852–3866 (2018).
118. A. Hariri et al., “Deep learning improves contrast in low-fluence photoacoustic imaging,” *Biomed. Opt. Express* **11**(6), 3360–3373 (2020).
119. J. Kerrick et al., “Novel deep learning architecture for optical fluence dependent photoacoustic target localization,” *Proc. SPIE* **10878**, 108781L (2019).
120. K. Johnstonbaugh et al., “A deep learning approach to photoacoustic wavefront localization in deep-tissue medium,” *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **67**, 2649–2659 (2020).
121. S. M. K. Ajith et al., “Deep learning-enhanced LED-based photoacoustic imaging,” *Proc. SPIE* **11240**, 1124038 (2020).
122. R. Manwar et al., “Deep learning protocol for improved photoacoustic brain imaging,” *J. Biophotonics* **13**(10), e202000212 (2020).
123. K. Hoffer-Hawlik and G. P. Luke, “AbsO2luteU-Net: tissue oxygenation calculation using photoacoustic imaging and convolutional neural networks,” ENGS 88 Honors Thesis (AB Students), p. 10, <https://digitalcommons.dartmouth.edu/engs88/10> (2019).
124. T. Chen et al., “A deep learning method based on U-Net for quantitative photoacoustic imaging,” *Proc. SPIE* **11240**, 112403V (2020).
125. C. Yang et al., “Quantitative photoacoustic blood oxygenation imaging using deep residual and recurrent neural network,” in *IEEE 16th Int. Symp. Biomed. Imaging*, pp. 741–744 (2019).
126. C. Yang and F. Gao, “EDA-net: dense aggregation of deep and shallow information achieves quantitative photoacoustic blood oxygenation imaging deep in human breast,” *Lect. Notes Comput. Sci.* **11764**, 246–254 (2019).
127. J. Gröhl et al., “Confidence estimation for machine learning-based quantitative photoacoustics,” *J. Imaging* **4**, 147 (2018).
128. O. H. Maghsoudi et al., “O-Net: an overall convolutional network for segmentation tasks,” in *Int. Workshop Mach. Learn. Med. Imaging*, pp. 199–209, Springer, Cham (2020).
129. C. Bench, A. Hauptmann, and B. T. Cox, “Towards accurate quantitative photoacoustic imaging: learning vascular blood oxygen saturation in 3D,” *J. Biomed. Opt.* **25**(8), 085003 (2020).

130. D. A. Durairaj et al., “Unsupervised deep learning approach for photoacoustic spectral unmixing,” *Proc. SPIE* **11240**, 112403H (2020).
131. J. Gröhl et al., “Estimation of blood oxygenation with learned spectral decoloring for quantitative photoacoustic imaging (LSD-qPAI),” arXiv:1902.05839 (2019).
132. I. Olefir et al., “Deep learning based spectral unmixing for optoacoustic imaging of tissue oxygen saturation,” *IEEE Trans. Med. Imaging* **39**, 3643–3654 (2020).
133. S. Narkhede, “Understanding AUC-ROC curve,” *Towards Data Science* 26 (2018).
134. Z. H. Zhou and X. Y. Liu, “Training cost-sensitive neural networks with methods addressing the class imbalance problem,” *IEEE Trans. Knowl. Data Eng.* **18**(1), 63–77 (2006).
135. M. Omar et al., “Ultrawideband reflection-mode optoacoustic mesoscopy,” *Opt. Lett.* **39**(13), 3911–3914 (2014).
136. K. He et al., “Deep residual learning for image recognition,” in *IEEE Conf. Comput. Vision and Pattern Recognit.*, pp. 770–778 (2016).
137. J. Aguirre et al., “Precision assessment of label-free psoriasis biomarkers with ultra-broadband optoacoustic mesoscopy,” *Nat. Biomed. Eng.* **1**(5), 0068 (2017).
138. Y. E. Boink et al., “A framework for directional and higher-order reconstruction in photoacoustic tomography,” *Phys. Med. Biol.* **63**(4), 045018 (2018).
139. E. Sidky, J. Jorgensen, and X. Pan, “Convex optimization problem prototyping for image reconstruction in computed tomography with the Chambolle–Pock algorithm,” *Phys. Med. Biol.* **57**, 3065–3091 (2012).
140. L. Zeune et al., “Combining contrast invariant L1 data fidelities with nonlinear spectral image decomposition,” *Lect. Notes Comput. Sci.* **10302**, 80–93 (2017).
141. A. Taruttis et al., “Motion clustering for deblurring multispectral optoacoustic tomography images of the mouse heart,” *J. Biomed. Opt.* **17**(1), 016009 (2012).
142. J. Xia et al., “Retrospective respiration-gated whole-body photoacoustic computed tomography of mice,” *J. Biomed. Opt.* **19**(1), 016003 (2014).
143. M. Schwarz et al., “Motion correction in optoacoustic mesoscopy,” *Sci. Rep.* **7**(1), 10386 (2017).
144. H. Zhao et al., “Motion correction in optical resolution photoacoustic microscopy,” *IEEE Trans. Med. Imaging* **38**(9), 2139–2150 (2019).
145. J. Staal et al., “Ridge-based vessel segmentation in color images of the retina,” *IEEE Trans. Med. Imaging* **23**(4), 501–509 (2004).
146. S. Ma, S. Yang, and H. Guo, “Limited-view photoacoustic imaging based on linear-array detection and filtered mean-backprojection-iterative reconstruction,” *J. Appl. Phys.* **106**, 123104 (2009).
147. M. Xu and L. Wang, “Analytic explanation of spatial resolution related to bandwidth and detector aperture size in thermoacoustic or photoacoustic reconstruction,” *Phys. Rev. E* **67**, 056605 (2003).
148. R. Manwar, K. Kratkiewicz, and K. Avnani, “Overview of ultrasound detection technologies for photoacoustic imaging,” *Micromachines* **11**(7), 692 (2020).
149. J. Hu, L. Shen, and G. Sun, “Squeeze-and-excitation networks,” in *IEEE/CVF Conf. Comput. Vision and Pattern Recognit.*, pp. 7132–7141 (2018).
150. G. Chen et al., “Rethinking the usage of batch normalization and dropout in the training of deep neural networks,” arXiv:1905.05928 (2019).
151. B. Lim et al., “Enhanced deep residual networks for single image super-resolution,” in *2017 IEEE Conf. Comput. Vision and Pattern Recognit. Workshops*, Honolulu, Hawaii, pp. 1132–1140 (2017).
152. M. M. Fraz et al., “An ensemble classification-based approach applied to retinal blood vessel segmentation,” *IEEE Trans. Biomed. Eng.* **59**(9), 2538–2548 (2012).
153. A. Hoover, V. Kouznetsova, and M. Goldbaum, “Locating blood vessels in retinal images by piecewise threshold probing of a matched filter response,” *IEEE Trans. Med. Imaging* **19**(3), 203–210 (2000).
154. M. Haltmeier et al., “A sparsification and reconstruction strategy for compressed sensing photoacoustic tomography,” *J. Acoust. Soc. Am.* **143**, 3838 (2018).
155. J. Schwab, S. Antholzer, and M. Haltmeier, “Learned backprojection for sparse and limited view photoacoustic tomography,” *Proc. SPIE* **10878**, 1087837 (2019).

156. P. Stefanov and Y. Yang, “Multiwave tomography in a closed domain: averaged sharp time reversal,” *Inverse Prob.* **31**, 065007 (2015).
157. A. Javaherian and S. Holman, “A continuous adjoint for photo-acoustic tomography of the brain,” *Inverse Prob.* **34**, 085003 (2018).
158. P. Stefanov and Y. Yang, “Multiwave tomography with reflectors: Landweber’s iteration,” *Inverse Prob. Imaging* **11**(2), 373–401 (2017).
159. S. Holan and J. Viator, “Automated wavelet denoising of photoacoustic signals for circulating melanoma cell detection and burn image reconstruction,” *Phys. Med. Biol.* **53**, N227–236 (2008).
160. Y. Gal and Z. Ghahramani, “Dropout as a Bayesian approximation: representing model uncertainty in deep learning,” in *Proc. of the 33rd Int. Conf. on Mach. Learn.*, Vol. **48**, pp. 1050–1059, PMLR (2016).
161. P. Isola et al., “Image-to-image translation with conditional adversarial networks,” in *IEEE Conf. Comput. Vision and Pattern Recognit.*, pp. 5967–5976 (2017).
162. G. Paltauf et al., “Experimental evaluation of reconstruction algorithms for limited view photoacoustic tomography with line detectors,” *Inverse Prob.* **23**(6), S81 (2007).
163. G. Paltauf et al., “Weight factors for limited angle photoacoustic tomography,” *Phys. Med. Biol.* **54**(11), 3303 (2009).
164. A. Alshaya et al., “Spatial resolution and contrast enhancement in photoacoustic imaging with filter delay multiply and sum beamforming technique,” in *IEEE Int. Ultrason. Symp.*, pp. 1–4 (2016).
165. G. Matrone et al., “The delay multiply and sum beamforming algorithm in ultrasound B-mode medical imaging,” *IEEE Trans. Med. Imaging* **34**(4), 940–949 (2015).
166. K. Cho et al., “Learning phrase representations using RNN encoder–decoder for statistical machine translation,” in *Proc. of the 2014 Conf. on Empirical Methods in Nat. Lang. Process.*, pp. 1724–1734 (2014).
167. T. Tong et al., “Image super-resolution using dense skip connections,” in *IEEE Int. Conf. Comput. Vision*, pp. 4809–4817 (2017).
168. X. Shi et al., “Convolutional LSTM Network: a machine learning approach for precipitation nowcasting,” in *Proc. 28th Int. Conf. Neural Inf. Process. Syst.*, MIT Press, Montreal, Canada, Vol. 1, pp. 802–810 (2015).
169. K. Tang et al., “Denoising method for photoacoustic microscopy using deep learning,” *Proc. SPIE* **11525**, 115252P (2020).
170. B. T. Cox et al., “Two-dimensional quantitative photoacoustic image reconstruction of absorption distributions in scattering media by use of a simple iterative method,” *Appl. Opt.* **45**(8), 1866–1875 (2006).
171. B. A. Kaplan et al., “Monte-Carlo-based inversion scheme for 3D quantitative photoacoustic tomography,” *Proc. SPIE* **10064**, 100645J (2017).
172. Q. B. Adam et al., “Quantitative photoacoustic imaging: correcting for heterogeneous light fluence distributions using diffuse optical tomography,” *J. Biomed. Opt.* **16**(9), 096016 (2011).
173. T. Kirchner, J. Gröhl, and L. Maier-Hein, “Context encoding enables machine learning-based quantitative photoacoustics,” *J. Biomed. Opt.* **23**(5), 056008 (2018).
174. Z. Zhou et al., “UNet++: a nested u-net architecture for medical image segmentation,” *Lect. Notes Comput. Sci.* **11045**, 3–11 (2018).
175. S. Tzoumas et al., “Eigenspectra optoacoustic tomography achieves quantitative blood oxygenation imaging deep in tissues,” *Nat. Commun.* **7**(1), 12121 (2016).
176. C. Bench, A. Hauptmann, and B. Cox, “Toward accurate quantitative photoacoustic imaging: learning vascular blood oxygen saturation in three dimensions,” *J. Biomed. Opt.* **25**(8), 085003 (2020).
177. A. B. Arrieta et al., “Explainable artificial intelligence (XAI): concepts, taxonomies, opportunities and challenges toward responsible AI,” *Inf. Fusion* **58**, 82–115 (2020).
178. F. N. Iandola et al., “SqueezeNet: AlexNet-level accuracy with 50× fewer parameters and <0.5 MB model size,” arXiv:1602.07360 (2016).
179. A. G. Howard et al., “MobileNets: efficient convolutional neural networks for mobile vision applications,” arXiv:1704.04861 (2017).

180. X. Zhang et al., “ShuffleNet: an extremely efficient convolutional neural network for mobile devices,” in *Proc. of the IEEE Conf. on Comput. Vision and Pattern Recognit.*, pp. 6848–6856 (2018).
181. J. Deng et al., “ImageNet: a large-scale hierarchical image database,” in *IEEE Conf. Comput. Vision and Pattern Recognit.*, pp. 248–255 (2009).
182. S. Walczak, I. Yegorova, and B. H. Andrews, “The effect of training set distributions for supervised learning artificial neural networks on classification accuracy,” in *Information Management: Support Systems and Multimedia Technology*, pp. 93–108, IGI Global (2003).
183. F. J. Pulgar et al., “On the impact of imbalanced data in convolutional neural networks performance,” *Lect. Notes Comput. Sci.* **10334**, 220–232 (2017).
184. X. Ma et al., “Multiple delay and sum with enveloping beamforming algorithm for photoacoustic imaging,” *IEEE Trans. Med. Imaging* **39**(6), 1812–1821 (2020).
185. M. Cao et al., “Spread spectrum photoacoustic tomography with image optimization,” *IEEE Trans. Biomed. Circuits Syst.* **11**(2), 411–419 (2017).
186. C. Rabut et al., “4D functional ultrasound imaging of whole-brain activity in rodents,” *Nat. Methods* **16**(10), 994–997 (2019).

Handi Deng graduated from the School of Precision Instrument and Opto-Electronics Engineering at Tianjin University in 2017. He is now studying for a doctorate in the Biophotonics Laboratory of the Department of Electronic Engineering at Tsinghua University. His research focuses on solving the problems of imperfect reconstruction and developing advanced photoacoustic computed tomography systems with clinical applications.

Hui Qiao received his BE degree in automation in 2013 and his PhD in control science and engineering from Tsinghua University, Beijing, China, in 2019. He is currently a postdoctoral scholar at the same university. His research interests include computational imaging, computational microscopy, and machine learning.

Qionghai Dai received his MS and PhD degrees in computer science and automation from Northeastern University, Shenyang, China, in 1994 and 1996, respectively. He is currently a full professor, the director of the School of Information Science and Technology, and the director of the Institute for Brain and Cognitive Sciences at Tsinghua University. He is also the chairman of Chinese Association for Artificial intelligence. His research centers on the interdisciplinary study of brain engineering and the next-generation artificial intelligence.

Cheng Ma received his BS degree in electronic engineering from Tsinghua University, Beijing, China, in 2004 and his PhD in electrical engineering from Virginia Tech, Blacksburg, Virginia, USA, in 2012. From 2012 to 2016, he was a postdoc in the Department of Biomedical Engineering at Washington University in St. Louis, St. Louis, Missouri, USA. He is now an associate professor in the Department of Electronic Engineering at Tsinghua University. His research interests include biophotonic imaging, in particular photoacoustic imaging.